الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.
CR No. 2055011643 س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC / 450398

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	**Sudhir Equipment Rental Company**			
Address	Al Jubail, Kingdom of Saudi Arabia.			
Description	**Pressure Relief Valve**			
Manufacturer	HARSHAD			
Part No.	9095799400			
Serial No.	SAC-250-299	**Receipt Date**	:	06-Jun-26
ID No.	N/A	**Calibration Date**	:	06-Jun-26
Location	Calibration Lab (Jubail)	**Calibration Due Date**	:	05-Dec-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No	Due Date
High Pressure Pneumatic Comparator	AME/25/TC-0711-A1	214E2408017	10-Jul-26
Digital Pressure Calibrator	AME/25/CC-0703-A2	2232209008	02-Jul-26

Calibrated By

UIC 68 CAL ENG

Calibration Engineer

Approved By

Quality/ Lab Manager

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement results shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515
www.ui.com.sa
ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ٣٤٤ ١٣ ٩٦٦+ / فاكس ٨٨٨١ ٣٤٤ ١٣ ٩٦٦+
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦+

CERTIFICATE NO: UIC / 450398

Customer	**Sudhir Equipment Rental Company**		
Address	Al Jubail, Kingdom of Saudi Arabia.		
Description	**Pressure Relief Valve**	**Calibration Date**	06-Jun-26
Manufacturer	HARSHAD	**Calibration Due Date**	05-Dec-26
Part No.	9.10E+09	**LTemperature**	24.2 °C
Serial No.	SAC-250-299	**Relative Humidity**	RH 51 %
ID No.	N/A		

Test Results

Set Pressure :	**8.0**	**BARG**	**Max Pressure**	**: 9.5 BARG**
Back Pressure:	**0.0**	**BARG**	**Size Inlet/Outlet**	**: 3/4" X 1"**
CDTP :	**8.0**	**BARG**		
Setting Accuracy : ± 3 % of Setting			**Type of Valve**	**: Metal Seated**
Pop Test : PASSED			**Test Medium**	**: Air**

BASIS OF TEST:

Ref Standard ASME Sec VIII,Div I

Summary of Test Procedure:

The Calibration of above relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic /pnenumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceabilty to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	Yes
The instrument was adjusted	No
The instrument was repaired	No

✦✦✦✦✦ **End of the Certificate** ✦✦✦✦✦

UIC 68 CAL ENGR

UNIVERSAL INSPECTION CO. LTD. C.R. 2055011643

Calibrated By: _____
Calibration Engineer

Approved by: _____
Quality/ Lab Manager