

CALIBRATION CERTIFICATE

DETAILS OF CALIBRATION CERTIFICATE

Certificate No	GTEC/AJ/26/NS18254-01	Date of Issue	23.07.2026
Date of Receipt	23.07.2026		
Calibration Date	23.07.2026		
Recommended Calibration Due	22.01.2027		

DETAILS OF CUSTOMER

Customer Name	M/s. Sudhir Equipment Rental Company.
Customer Address	Dammam, Kingdom of Saudi Arabia.

Glory Technology Calibration & Inspection Company (GTEC) certifies that the below described instrument has been calibrated and the measurements are traceable through an unbroken chain to National/International Standards and International System of Units (SI). This calibration certificate documents the traceability of the reference standards used and are available upon the request of customer. The calibration results are documented on the succeeding page/pages of this certificate. The customer is obliged to have the instruments calibrated at regular intervals.

DETAILS OF UNIT UNDER CALIBRATION (UUC)

Description	Pressure Relief Valve
Set Pressure	17 bar
Make / Model	--
Serial No / Identification No	SAC-435-019
Physical Condition	Satisfactory

DETAILS OF REFERENCE EQUIPMENTS/STANDARDS

Description	Precision Digital Pressure Gauge
Model / Serial No	CPG 1500 / 1A00EU85VIS
Certificate No / Validity	BKGTS/MIN2527.04/2025 / 28.10.2026
Traceability	BKGTS / SAAC N-C-00552

DETAILS OF ENVIRONMENTAL CONDITIONS

Temperature	22.7°C	Relative Humidity	43.2%RH
-------------	--------	-------------------	---------

DETAILS OF CALIBRATION PROCEDURE

Reference Procedure / Standard	GTEC/WI/013 / API 527
--------------------------------	-----------------------

Calibrated By



Sangeetha Sabu

Calibration Engineer



Authorized By

C. Balasubramanian

C. Balasubramanian

Technical Manager



Page 1 of 2

www.gtecgulf.com

Certificate No ,	GTEC/AJ/26/NS18254-01
------------------	-----------------------

CALIBRATION RESULTS

Test Result

Set Pressure	17 bar	Test Medium	Air
CDTP	17 bar	Valve Inlet Size	3/4"
Start to Leak Pressure	17 bar	Valve Outlet Size	1"
Seat Leak Test Pressure	15 bar	Pressure Setting Tolerance	± 3 %

Physical Inspection Result

Sl. No	Description	Condition	Remarks
1	Valve Body	Good	-
2	Bonnet	Good	-
3	Cap	Good	-
4	Stem	Good	-
5	Adjusting Screw	Good	-
6	Seat	Good	-

Note.

1. The UUC was calibrated at Lab.
2. Calibration Certificate shall not be reproduced, except in full, without written approval of GTEC.
3. Results reported are valid at the time and under stated environmental conditions of calibration.
4. This Certificate is relates only to the particular item calibrated.

*****End of Certificate*****

Calibrated By



Sangeeth Sabu

Calibration Engineer



www.gtecgulf.com



Authorized By

C. Balasubramanian

Technical Manager

Page 2 of 2