



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/RIY/48219

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Sudhir Equipment Rental Company.**
Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**

Manufacturer	HARSHAD	Receipt Date	:	22-Sep-26
Part No.	9095799400	Calibration Date	:	22-Sep-26
Serial No.	2404806	Report Date	:	22-Sep-26
ID No.	SAC-250-214	Issue date	:	22-Sep-26
Location	Calibration Lab	Calibration Due Date	:	21-Mar-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements


Reference Standard Equipment Used

Description	Model/Cert No	Serial No	Due Date
High Pressure Pneumatic Calibrator	SA178189-223867	2011H13013	23-Nov-26
Digital Pressure Calibrator	SA178189-223866	2.2332E+11	26-Nov-26

Calibrated By


Calibration Engineer

Approved By


QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١
هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO : UIC/RIY/48219

PAGE

1 of 2

Customer	Sudhir Equipment Rental Company.		
Address	Dammam, Kingdom of Saudi Arabia		
Description	Pressure Relief Valve	Calibration Date	22-Sep-26
Manufacturer	HARSHAD	Calibration Due Date	21-Mar-27
Model	9095799400	Temperature	23.0 ± 2°C
Serial No.	2404806	Relative Humidity	RH 45% to 75%
ID No.	SAC-250-214		

Test Results

Set Pressure :	9.5	BAR	Spring Range	: N/A
Back Pressure:	0.0	BAR	Size Inlet / Outlet	: 3/4"
CDTP :	9.5	BAR	Type of Valve	: Metal Seated
Setting Accuracy : ± 3 % of Setting			Test Medium	: Air/Water/Oil/Gas
Pop Test : PASSED				

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company Maintains traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	No
The instrument was adjusted	Yes
The instrument was repaired	No

Result:

Calibration Result is pass.



Calibrated By: _____
Calibration Engineer



Approved by:
QA/QC Engineer



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/RIY/48220

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Sudhir Equipment Rental Company.**
Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Gauge**

Manufacturer VEETHREE

Model N/A

Serial No. N/A

ID. No. SAC-250-214

Location Calibration Lab

Receipt Date : 22-Sep-26

Calibration Date : 22-Sep-26

Report Date : 22-Sep-26

Issue date : 22-Sep-26

Calibration Due Date : 21-Mar-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model/Cert No	Serial No	Due Date
High Pressure Pneumatic Calibrator	SA178189-223867	2011H13013	23-Nov-26
Digital Pressure Calibrator	SA178189-223866	2.2332E+11	26-Nov-26

Calibrated By

13
Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO: UIC/RIY/48220

Customer
Address

Sudhir Equipment Rental Company.
Dammam, Kingdom of Saudi Arabia

Description
Manufacturer
Model
Serial No.
ID. No.

Pressure Gauge
VEETHREE
N/A
N/A
SAC-250-214

Calibration Date
Calibration Due Date
Temperature
Relative Humidity

22-Sep-26
21-Mar-27
23.0 ± 2 °C
RH 45% to 75%

CALIBRATION RESULTS FOR PRESSURE GAUGE

Min. Range: 0.0 (Kglcm2) **Max. Range:** 16.0 (Kglcm2)

Function Tested		Gauge Pressure	
UUT Setting (Kglcm2)	Actual reading in Master (Kglcm2)	Error (Kglcm2)	Error (% of Range) (FS)
0.00	0.00	0.00	0.000
2.00	2.10	0.10	0.625
6.00	6.20	0.20	1.250
12.00	12.10	0.10	0.625
16.00	16.10	0.10	0.625

Calibration Procedure:

UIC/P/PPGCP Rev. 02

Reference Standard:

As Generally given in MSL Technical guide & Tolerance as per EN 837 standard & based on the Manufacture's Specifications.

Remarks / Deviations IF ANY:

The results stated in this certificate relate only to the item calibrated.
The item was received in good condition.


Result:

Calibration Result is pass.

Calibrated By: 
Calibration Engineer

***** End of the Certificate *****

UIC-QP-7/8-2-17,REV 02

Approved by: 
QA/QC Engineer