الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.
CR No. 2055011643
س. ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/ALH/53548

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	**Sudhir Equipment Rental Company**			
Address	Al Hasa, Kingdom Of Saudi Arabia.			
Description	**Pressure Relief Valve**			
Manufacturer	ATLAS COPCO			
Part No	N/A			
Serial No.	SAC-500-112			
ID No.	N/A	Calibration Date	:	23-Aug-26
Location	PS-1	Calibration Due Date	:	22-Feb-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Cert. No./Model	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA178189-227063	214E2207012	27-Apr-27
Digital Pressure Calibrator	SA178189-227064	223210304	27-Apr-27

Calibrated By

Calibration Engineer

QA/QC

UIC-QP-7.8-24 F02, REV.0

Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515
www.ui.com.sa
ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ٣٤٤ ١٣ ٩٦٦+ / فاكس ٨٨٨١ ٣٤٤ ١٣ ٩٦٦+
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦+

CERTIFICATE NO: UIC/ALH/53548

Customer	**Sudhir Equipment Rental Company**		
Address	Al Hasa, Kingdom Of Saudi Arabia.		
Description	**Pressure Relief Valve**	**Calibration Date**	23-Aug-26
Manufacturer	ATLAS COPCO	**Calibration Due Date**	22-Feb-27
Part No	N/A	**Temperature**	25.2°C
Serial No.	SAC-500-112	**Relative Humidity**	RH 48%
ID No.	N/A		

Test Results

Set Pressure	: 14.0	BAR	Spring Range : N/A
Back Pressure	: 0.00	BAR	Size Inlet : 1"
CDTP	: 14.0	BAR	
Setting Accuracy	: ± 3 % of Setting		Type of Valve : Metal Seated
Pop Test : PASSED			Test Medium : Air

BASIS OF TEST:

Ref Standard ASME Sec VIII,Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pnenumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceabilty to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	Yes
The instrument was adjusted	No
The instrument was repaired	No

* * * * * End of the Certificate * * * * *

UIC 34 CAL-ENGR

QA/QC UNIVERSAL INSPECTION CO. LTD. الشركة العالمية لخدمات التفتيش المحدودة

Calibrated By: _____
Calibration Engineer

Approved by: _____
QA/QC Engineer