



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No.2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/RIY/39221

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Sudhir Equipment Rental Company.**
Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**

Manufacturer	HARSHAD	Receipt Date	:	22-Jan-26
Part No.	9095799400	Calibration Date	:	23-Jan-26
Serial No.	2404883	Report Date	:	23-Jan-26
ID No.	SAC-250-212	Issue date	:	23-Jan-26
Location	Calibration Lab	Calibration Due Date	:	22-Jul-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements


Reference Standard Equipment Used

Description	Model/Cert No	Serial No	Due Date
Water Pressure Comparator	VIC-25-42247	990S2412008	04-May-26
Digital Pressure Gauge	VIC-25-57095	2232412038	04-May-26

Calibrated By

Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦

CERTIFICATE NO : UIC/RIY/39221

PAGE

1 of 2

Customer
Address**Sudhir Equipment Rental Company.**
Dammam, Kingdom of Saudi ArabiaDescription
Manufacturer
Model
Serial No.
ID No.**Pressure Relief Valve**
HARSHAD
9095799400
2404883
SAC-250-212Calibration Date 23-Jan-26
Calibration Due Date 22-Jul-26
Temperature 23.0 ± 2°C
Relative Humidity RH 45% to 75%Test ResultsSet Pressure : 9.5 BAR
Back Pressure: 0.0 BAR
CDTP : 9.5 BAR

Spring Range : N/A

Size Inlet / Outlet : 3/4"

Setting Accuracy : ± 3 % of Setting

Type of Valve : Metal Seated

Pop Test : PASSED

Test Medium : Air/Water/Oil/Gas

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company Maintains traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)

No

The instrument was adjusted

Yes

The instrument was repaired

No

Result:

Calibration Result is pass.



++++ End of the Certificate +++++

Calibrated By:



Calibration Engineer

UIC-QP-7.8.2-17, REV 02

Approved by:

QA/QC Engineer