



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/47852

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company		
Location	Dammam, Kingdom of Saudi Arabia		
Description	Pressure Gauge		
Manufacturer	Pricol		
Model	N/A		
Serial No.	PG-47852	Receipt Date	12-Jan-25
ID.No	SAC-250-229	Calibration Date	12-Jan-25
Location	Calibration Lab (Dammam)	Calibration Due Date	11-Jul-25

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Cert. No./Model	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA 178189-210065	214E191209	23-Apr-25
Digital Pressure Calibrator	SA 178189-210066	23162191108	23-Apr-25


Calibrated By


Calibration Engineer

UIC-QP-7.8-24 F02, REV 0



Approved By


QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١
هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

Customer **Sudhir Equipment Rental Company**
Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Gauge**

Manufacturer Pricol

Model N/A

Serial No. PG-47852

ID.No SAC-250-229

Calibration Date 12-Jan-25

Calibration Due Date 11-Jul-25

Temperature Range 23.3°C

Relative Humidity RH 48%

Calibration Test Results for Pressure Gauge

Min. Range: 0.0 kgf/cm²

Max. Range: 16.0 kgf/cm²

Function Tested		Gauge Pressure	
UUT Setting kgf/cm ²	Actual Reading in Master kgf/cm ²	Error kgf/cm ²	Error (FS) (% of Range) kgf/cm ²
0.00	0.00	0.000	0.000
2.00	2.10	0.100	0.625
4.00	4.08	0.080	0.500
8.00	8.05	0.050	0.313
14.00	14.10	0.100	0.625
16.00	16.09	0.090	0.562

Calibration Procedure:

UIC/P/PPGCP Rev. 00

Reference Standard:

As Generally given in MSL Technical guide & Tolerance as per EN 837 standard & based on the Manufacture's Specifications.

Remarks / Deviations IF ANY:

The results stated in this certificate relate only to the item calibrated.
 The item was received in good condition.



***** End of the Certificate *****



Calibrated By: _____
 Calibration Engineer

UIC-QP-7.8-24 F02, REV 0

Approved by: _____
 QA/QC Engineer