



الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/RIY/29071

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Rawabi Butec Co.Ltd.**

Address Al-Khobar, Kingdom of Saudi Arabia.

Description **Pressure Relief Valve**

Manufacturer	HARSHAD	Calibration Date	: 6-Apr-25
Part No.	9095799400	Report Date	: 6-Apr-25
Serial No.	2404867	Issue date	: 6-Apr-25
Location	Calibration Lab	Calibration Due Date	: 5-Oct-25

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model/Cert No	Serial No	Due Date
High Pressure Pneumatic Calibrator	VIC-24-41692	2011H13013	15-Sep-25
Digital Pressure Calibrator	VIC-24-56304	223320200723	23-Nov-25

Calibrated by



Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦

هاتف ٠٥١٥ ٣٦٧ ٩٦٦

CERTIFICATE NO : UIC/RIY/29071

Customer Address **Rawabi Butec Co.Ltd.**
Al-Khobar, Kingdom of Saudi Arabia.

Description	Pressure Relief Valve	Calibration Date	6-Apr-25
Manufacturer	HARSHAD	Calibration Due Date	5-Oct-25
Part No.	9095799400	Temperature	23.0 ± 2°C
Serial No.	2404867	Relative Humidity	RH 45% to 75%

Test Results

Set Pressure :	8.0	BAR	Spring Range	: N/A
Back Pressure:	0.0	BAR	Size	: 3/4"
CDTP :	8.0	BAR	Type of Valve	: Metal Seated
Setting Accuracy : ± 3 % of Setting			Test Medium	: Air
Pop Test : PASSED				

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div i

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company Maintains traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)

No

The instrument was adjusted

No

The instrument was repaired

No

++++ End of the certificate +++++

Calibrated By: _____

Calibration Engineer



Approved by: _____

QA/QC Engineer