



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC / 362808

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Sudhir Equipment Rental Company**
Address Dammam, Kingdom of Saudi Arabia.

Description **Pressure Relief Valve**

Manufacturer HARSHAD

Model/Part No. 9095799400

Serial No. 2313560

ID No. N/A

Location Calibration Lab (Jubail)

Receipt Date : 22-Jun-25

Calibration Date : 24-Jun-25

Calibration Due Date : 23-Dec-25

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Cert. No./Model	Serial No	Due Date
High Pressure Pneumatic Comparator	SA 178189-216695	214E2207005	19-Dec-25
Digital Pressure Calibrator	2024005367	710190419	14-Jul-25

Calibrated By



Calibration Engineer



Approved By

Quality / Lab Manager

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO: UIC / 362808

Customer
Address**Sudhir Equipment Rental Company**
Dammam, Kingdom of Saudi Arabia.Description
Manufacturer
Model/Part No.
Serial No.
ID No.**Pressure Relief Valve**
HARSHAD
9095799400
2313560
N/ACalibration Date
Calibration Due Date
Lab Temperature
Relative Humidity
24-Jun-25
23-Dec-25
24.2 °C
RH 50 %Test ResultsSet Pressure : 5.00 BARG
Back Pressure: 0.00 BARG
CDTP : 5.00 BARG
Setting Accuracy : ± 3 % of Setting
Pop Test : PASSEDSpring Range : N/A
Size Inlet : 3/4"
Type of Valve : Metal Seated
Test Medium : AirBASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic /pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of InstrumentWith in the tolerance on receipt(as found)
The instrument was adjusted
The instrument was repairedNo
No
No

***** End of the Certificate *****

Calibrated By:

Calibration Engineer

UIC-QP-7.8-24 F02, REV 0

Approved by:

Quality Lab Manager