



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

AIR QUALITY TEST REPORT

REPORT NO: UIC/AIR/2026/0144

DATE OF REPORT: 18 JAN 2026

DATE OF TESTING: 17 JAN 2026

INTRODUCTION:

A cylinder containing breathing air was received from **M/S SINOPEC-5**, on 17th JAN 2026. It was requested by the client to carryout purity testing as per CGA Standard Grade "D" breathing air.

Customer Name : SINOPEC-5
Location : ABUHADRIYAH, KSA.
Sample Identification : AIR COMPRESSOR
Sticker Number : AQT/00571
Make : ATLAS COPCO
Model : XAS 500-175
SL.No : PE2460291
ID.No : SAC-500-218

TEST PROCEDURE:

The major components of the air sample were determined by using gas chromatograph equipped with thermal conductivity and flame ionization detectors. The other impurities are analyzed by Chemical reaction method using detection tubes.

SPECIFICATIONS:

CGA Commodity Specification for Air, ANSI/CGA G-7.1-2018

RESULTS


Test Description	Unit	Results	CGA Grade - D Limits (Max)
Oxygen (O ₂)	(Vol. %)	20.22	19.5 – 23.5
Carbon Dioxide (CO ₂)	(Vol. %)	0.05	0.10
Carbon Monoxide (CO)	(ppm)	< 2.0	10.0
Oil Condensed	(mg/m ³)	< 2.0	< 5.0
Moisture Content	(ppm)	61	67.0
Odour	---	None	None

REMARKS:

The above test result of the air sample complies with the Compressed Gas Association Standard for Grade "D" Breathing Air for SCBA or Airline respirators. Further precautionary measure time to time, we would recommend to clean the filter tap prior to use in breathing purposes.

Tested By:

Materials Analysis Dept.



UIC-AQT-RT-01/REV.NO-00

Approved By:

QA/QC Manager

This report relates only to the sample tested and only shall be reproduced in full with written approval of UIC testing laboratory

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥



الشركة العالمية للخدمات التقنية المحدودة
UNIVERSAL INSPECTION CO. LTD.

TESTING DIVISION

AIR QUALITY TESTING FOR AIR COMPRESSOR

NO.AQT/00571

CERTIFICATE NO : UC-AIR.2026-0144
EQUIPMENT ID NO : SAC-D00-218
SERIAL NUMBER : PE2460291
SAMPLE COLLECTION DATE : 17. JAN. 2026
REPORT DATE : 18. JAN. 2026
TESTING REPORT : PASSED
TESTED BY : M. Rajkumar

Contact No. Tel.: +966 13 344 8882, Fax: +966 13 344 8881

Website : www.ui.com.sa

DOC.NO UIC-AQT-RT-01