



الشركة العالمية لخدمات التفتيش المحدودة

UNIVERSAL INSPECTION CO. LTD.

CR No.2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/51393

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Sudhir Equipment Rental Company**

Location Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**

Manufacturer Harshad

Part No. 0832100066

Serial No. 2424221

ID. No. SAC-500-308

Location Calibration Lab (Dammam)

Receipt Date : 02-Mar-25

Calibration Date : 02-Mar-25

Calibration Due Date : 01-Sep-25

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA 178189-210065	214E191209	23-Apr-25
Digital Pressure Calibrator	SA 178189-210066	23162191108	23-Apr-25

Calibrated By

Calibration Engineer

UIC-OP-7.8-24 F02, REV 0

Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO: UIC/DAM/51393

Page

2 of 2

Customer **Sudhir Equipment Rental Company**
Location Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**
Manufacturer Harshad
Model 0832100066
Serial No. 2424221
ID. No. SAC-500-308

Calibration Date 02-Mar-25
Calibration Due Date 01-Sep-25
Temperature Range 23.2°C
Relative Humidity RH 49%

Test Results

Set Pressure : 12.00 BAR Spring Range : N/A
 Back Pressure : 0.00 BAR Size Inlet/Outlet : 3/4"
 CDTP : 12.00 BAR Type of Valve : Metal Seated
 Setting Accuracy : $\pm 3\%$ of Setting Test Medium : Air
 Pop Test : PASSED

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found) No
 The instrument was adjusted Yes
 The instrument was repaired No

++++ End of the Certificate +++++



Calibrated By:
 Calibration Engineer



UIC-QP-7.8-24 F02, REV 0

Approved by:
 QA/QC Engineer