



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/RIY/37986

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Sudhir Equipment Rental Company.**
Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Gauge**
Manufacturer N/A
Model N/A
Serial No. N/A
ID. No. SAC-500-394
Location Calibration Lab

Receipt Date : 18-Jan-26
Calibration Date : 18-Jan-26
Report Date : 18-Jan-26
Issue date : 18-Jan-26
Calibration Due Date : 17-Jul-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model/Cert No	Serial No	Due Date
High Pressure Pneumatic Calibrator	SA178189-223867	2011H13013	23-Nov-26
Digital Pressure Calibrator	SA178189-223866	223320200723	26-Nov-26

Calibrated By
UIC
13
CAL ENGR
Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦
هاتف ٠٥١٥ ١٣ ٣٦٧ ٩٦٦

CERTIFICATE NO: UIC/RIY/37986

Customer **Sudhir Equipment Rental Company.**

Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Gauge**

Manufacturer N/A

Model N/A

Serial No. N/A

ID. No. SAC-500-394

Calibration Date 18-Jan-26

Calibration Due Date 17-Jul-26

Temperature 23.0 ± 2 °C

Relative Humidity RH 45% to 75%

CALIBRATION RESULTS FOR PRESSURE GAUGE

Min. Range: 0.0 (BAR)

Max. Range: 25.0 (BAR)

Function Tested		Gauge Pressure	
UUT Setting (BAR)	Actual reading in Master (BAR)	Error (BAR)	Error (% of Range) (FS)
0.00	0.00	0.00	0.000
5.00	4.98	-0.02	-0.080
15.00	14.99	-0.01	-0.040
20.00	20.12	0.12	0.480
25.00	25.30	0.30	1.200

Calibration Procedure:

UIC/P/PPGCP Rev. 02

Reference Standard:

As Generally given in MSL Technical guide & Tolerance as per EN 837 standard & based on the Manufacture's Specifications.

Remarks / Deviations IF ANY:

The results stated in this certificate relate only to the item calibrated.
The item was received in good condition.



++++ End of the Certificate +++++

Calibrated By:



Calibration Engineer

UIC-QP-7.8-2-17,REV 02

Approved by:

QA/QC Engineer