

## CALIBRATION CERTIFICATE

### DETAILS OF CALIBRATION CERTIFICATE

|                             |                       |               |            |
|-----------------------------|-----------------------|---------------|------------|
| Certificate No              | GTEC/AJ/25/NS14310-01 | Date of Issue | 17.07.2025 |
| Date of Receipt             | 17.07.2025            |               |            |
| Calibration Date            | 17.07.2025            |               |            |
| Recommended Calibration Due | 16.01.2026            |               |            |

### DETAILS OF CUSTOMER

|                  |                                       |
|------------------|---------------------------------------|
| Customer Name    | M/s. Sudhir Equipment Rental Company. |
| Customer Address | Dammam, Kingdom of Saudi Arabia.      |

Glory Technology Calibration & Inspection Company (GTEC) certifies that the below described instrument has been calibrated, complying with the National/International Standards. This calibration certificate documents the traceability of the reference standards used and are available upon the request of customer. The calibration results are documented on the succeeding page/pages of this certificate. The customer is obliged to have the instruments calibrated at regular intervals.

### DETAILS OF UNIT UNDER CALIBRATION (UUC)

|                               |                       |
|-------------------------------|-----------------------|
| Description                   | Pressure Relief Valve |
| Spring Range / Set Pressure   | 16 bar / 16 bar       |
| Make / Model                  | HARSHAD               |
| Serial No / Identification No | SAC-500-395           |
| Physical Condition            | Satisfactory          |

### DETAILS OF REFERENCE EQUIPMENTS/STANDARDS

|                           |                                 |
|---------------------------|---------------------------------|
| Description               | Digital Pressure Gauge          |
| Model / Serial No         | CPG 1500 / 1A02N5D8M2C          |
| Certificate No / Validity | MTE-CC-J80031077-3 / 29.10.2025 |
| Traceability              | MTE / IAS CL-248                |

### DETAILS OF ENVIRONMENTAL CONDITIONS

|             |        |                   |         |
|-------------|--------|-------------------|---------|
| Temperature | 23.8°C | Relative Humidity | 42.3%RH |
|-------------|--------|-------------------|---------|

### DETAILS OF CALIBRATION PROCEDURE

|                                |                       |
|--------------------------------|-----------------------|
| Reference Procedure / Standard | GTEC/WI/013 / API 527 |
|--------------------------------|-----------------------|

Calibrated By



Abhijith V N

Calibration Engineer



www.gteccalibration.com



Authorized By

C. Balasubramanian

Technical Manager

Page 1 of 2

Certificate No : GTEC/AJ/25/NS14310-01

### CALIBRATION RESULTS

#### Test Result

|                         |        |                            |       |
|-------------------------|--------|----------------------------|-------|
| Set Pressure            | 16 bar | Test Medium                | Water |
| CDTP                    | 16 bar | Valve Inlet Size           | 1"    |
| Start to Leak Pressure  | 16 bar | Valve Outlet Size          | 1.5"  |
| Seat Leak Test Pressure | 14 bar | Pressure Setting Tolerance | ± 3 % |

#### Physical Inspection Result

| Sl. No | Description     | Condition | Remarks |
|--------|-----------------|-----------|---------|
| 1      | Valve Body      | Good      | -       |
| 2      | Bonnet          | Good      | -       |
| 3      | Cap             | Good      | -       |
| 4      | Stem            | Good      | -       |
| 5      | Adjusting Screw | Good      | -       |
| 6      | Nozzle          | Good      | -       |

#### Note:

1. The UUC was calibrated at Lab.
2. Calibration Certificate shall not be reproduced, expect in full, without written approval of GTEC.
3. Results reported are valid at the time and under stated environmental conditions of calibration.
4. This Certificate is relates only to the particular item calibrated.

\*\*\*\*\*End of Certificate\*\*\*\*\*

Calibrated By



Abhijith V N

Calibration Engineer



www.gteccalibration.com



Authorized By

*C. Balasubramanian*  
C. Balasubramanian

Technical Manager

Page 2 of 2