



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/ALH/40278

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer Sudhir Equipment Rental Company.

Address Gosp-1, Kingdom Of Saudi Arabia.

Description Pressure Relief Valve

Manufacturer ATLAS COPCO

Model N/A

Serial No. SAC-500-431

Id No. N/A

Calibration Date : 18-Mar-26

Location Calibration Lab (Al Hasa)

Calibration Due Date : 17-Sep-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No.	Due Date
High Pressure Pneumatic Pressure Source	SA17-8189-219407	214E2207012	21-Apr-26
Digital Pressure Calibrator	SA17-8189-219408	223210304	21-Apr-26


Calibrated By



Calibration Engineer

UIC-QP-7-8-24-F02, REV 0

Approved By



QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٨٨٨٢ ٣٤٤ ٩٦٦ / فاكس ٨٨٨١ ٣٤٤ ٩٦٦

هاتف ٥١٥ ٣٦٧ ٩٦٦

CERTIFICATE NO: UIC/ALH/40278

Page

2 of 2

Customer **Sudhir Equipment Rental Company.****Address** Gosp-1, Kingdom Of Saudi Arabia.**Description** **Pressure Relief Valve****Manufacturer** ATLAS COPCO**Model** N/A**Serial No.** SAC-500-431**Calibration Date** 18-Mar-26**Calibration Due Date** 17-Sep-26**Temperature** 25.2°C**Relative Humidity** RH 48%Test Results

Set Pressure : 14.0 BAR

Spring Range : N/A

Back Pressure : 0.00 BAR

Size Inlet / Outlet : 1" X 1 ¼"

CDTP : 14.0 BAR

Setting Accuracy : $\pm 3\%$ of Setting

Type of Valve : Metal Seated

Pop Test : PASSED

Test Medium : Air

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)

Yes

The instrument was adjusted

No

The instrument was repaired

No

++++ End of the Certificate +++++

Calibrated By: _____



Calibration Engineer



UIC-QP-7.8.2-F02, REV 0

Approved by: _____

QA/QC Engineer