



الشركة العالمية لخدمات التفتيش المحدودة

UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

AIR QUALITY TEST REPORT

REPORT NO: UIC/AIR/2025/2069

DATE OF REPORT: 23 DEC 2025

DATE OF TESTING: 23 DEC 2025

INTRODUCTION:

A cylinder containing breathing air was received from M/S SUDHIR EQUIPMENT RENTAL COMPANY, on 22nd DEC 2025. it was requested by the client to carryout purity testing as per CGA Standard Grade "D" breathing air.

Customer Name : SUDHIR EQUIPMENT RENTAL COMPANY,
Location : DAMMAM, KSA.
Sample Identification : AIR COMPRESSOR
Sticker Number : AQT/00320
Make : ATLASCOPCO
Model : XAS 500-175
SL.No : PNE0003474
ID.No : SAC-500-433.

TEST PROCEDURE:

The major components of the air sample were determined by using gas chromatograph equipped with thermal conductivity and flame ionization detectors. The other impurities are analyzed by Chemical reaction method using detection tubes.

SPECIFICATIONS:

CGA Commodity Specification for Air, ANSI/CGA G-7.1-2018

RESULTS

Test Description	Unit	Results	CGA Grade - D Limits (Max)
Oxygen (O ₂)	(Vol. %)	20.91	19.5 – 23.5
Carbon Dioxide (CO ₂)	(Vol. %)	0.04	0.10
Carbon Monoxide (CO)	(ppm)	< 1.0	10.0
Oil Condensed	(mg/m ³)	< 2.0	< 5.0
Moisture Content	(ppm)	61	67.0
Odour	---	None	None

REMARKS:

The above test result of the air sample complies with the Compressed Gas Association Standard for Grade "D" Breathing Air for SCBA or Airline respirators. Further precautionary measure time to time, we would recommend to clean the filter tap prior to use in breathing purposes.

Tested By:

Materials Analysis Dept.

UIC-AQT-RT-01/REV. NO:00

Approved By:

QA/QC Manager

This report relates only to the sample tested and only shall be reproduced in full with written approval of UIC testing laboratory

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥