



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No.2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC / SAF / 69629

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company		
Address	Dammam, Kingdom of Saudi Arabia		
Description	Pressure Relief Valve		
Manufacturer	Atlas Copco		
Type.	AK28X-50T		
Serial No.	2307822	Receipt Date	: 03-Nov-24
ID No.	SAC-825-173	Calibration Date	: 03-Nov-24
Location	Calibration Lab (Safaniya)	Calibration Due Date	: 02-May-25

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements


Reference Standard Equipment Used

Description	Cert. No./Model	Serial No	Due Date
High Pressure Pneumatic Comparator	VIC-24-41658	214E191210	10-Aug-25
Digital Pressure Gauge	178189-213567	223162191111	01-Sep-25

Calibrated By

32

Calibration Engineer



Approved By

[Signature]

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦
هاتف ٠٥١٥ ١٣ ٩٦٦

CERTIFICATE NO:		UIC / SAF / 69629		PAGE	2 of 2
Customer Address	Sudhir Equipment Rental Company Dammam, Kingdom of Saudi Arabia				
Description	Pressure Relief Valve				
Manufacturer	Atlas Copco	Calibration Date	03-Nov-24		
Model	AK28X-50T	Calibration Due Date	02-May-25		
Serial No.	2307822	Temperature	23.0 ± 2°C		
ID No.	SAC-825-173	Relative Humidity	RH 45% to 75%		
Test Results					
Set Pressure :	16.0 BAR	Spring Range :	N/A		
Back Pressure:	0.0 BAR	Size Inlet /Outlet :	1" X 1 ½"		
CDTP :	16.0 BAR	Type of Valve :	Metal Seated		
Setting Accuracy :	+ 3 % of Setting		Test Medium :	Air	
Pop Test :	PASSED				
BASIS OF TEST:					
Ref Standard ASME Sec VIII, Div I					
Summary of Test Procedure:					
The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.					
Traceability:					
The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.					
Condition of Instrument					
With in the tolerance on receipt(as found)	No				
The instrument was adjusted	Yes				
The instrument was repaired	No				
+++++ End of the Certificate +++++					
Calibrated By:					
	Calibration Engineer		 Approved by: QA/QC Engineer		
UIC-QP-7.8-24 F02, REV 0					