

CALIBRATION CERTIFICATE

DETAILS OF CALIBRATION CERTIFICATE

Certificate No	GTEC/AJ/26/NS14361-01	Date of Issue	14.06.2026
Date of Receipt	14.06.2026		
Calibration Date	14.06.2026		
Recommended Calibration Due	13.12.2026		

DETAILS OF CUSTOMER

Customer Name	M/s. Sudhir Equipment Rental Company.
Customer Address	Dammam, Kingdom of Saudi Arabia.

Glory Technology Calibration & Inspection Company (GTEC) certifies that the below described instrument has been calibrated and the measurements are traceable through an unbroken chain to National/International Standards and International System of Units (SI). This calibration certificate documents the traceability of the reference standards used and are available upon the request of customer. The calibration results are documented on the succeeding page/pages of this certificate. The customer is obliged to have the instruments calibrated at regular intervals.

DETAILS OF UNIT UNDER CALIBRATION (UUC)

Description	Pressure Relief Valve
Spring Range / Set Pressure	16 bar / 12 bar
Make / Model	ATLAS COPCO
Serial No / Identification No	2307827 / SAC.825-178
Physical Condition	Satisfactory

DETAILS OF REFERENCE EQUIPMENTS/STANDARDS

Description	Precision Digital Pressure Gauge
Model / Serial No	CPG 1500 / 1A00EU85VIS
Certificate No / Validity	BKGTS/MIN2527.04/2025 / 28.10.2026
Traceability	BKGTS / SAAC N-C-00552

DETAILS OF ENVIRONMENTAL CONDITIONS

Temperature	22.7°C	Relative Humidity	43.2%RH
-------------	--------	-------------------	---------

DETAILS OF CALIBRATION PROCEDURE

Reference Procedure / Standard	GTEC/WI/013 / API 527
--------------------------------	-----------------------

Calibrated By



Sangeeth Sabu

Calibration Engineer



www.gtec Calibration.com



Authorized By

C. Balasubramanian

Technical Manager

Page 1 of 2

Certificate No .	GTEC/AJ/26/NS14361-01
------------------	-----------------------

CALIBRATION RESULTS

Test Result

Set Pressure	12 bar	Test Medium	Air
CDTP	12 bar	Valve Inlet Size	1"
Start to Leak Pressure	12 bar	Valve Outlet Size	1 1/2"
Seat Leak Test Pressure	11 bar	Pressure Setting Tolerance	± 3 %

Physical Inspection Result

Sl. No	Description	Condition	Remarks
1	Valve Body	Good	-
2	Bonnet	Good	-
3	Cap	Good	-
4	Stem	Good	-
5	Adjusting Screw	Good	-
6	Seat	Good	-

Note.

1. The UUC was calibrated at Lab.
2. Calibration Certificate shall not be reproduced, except in full, without written approval of GTEC.
3. Results reported are valid at the time and under stated environmental conditions of calibration.
4. This Certificate is relates only to the particular item calibrated.

*****End of Certificate*****

Calibrated By



Sangeeth Sabu

Calibration Engineer



www.gteccalibration.com

Authorized By

C. Balasubramanian

Technical Manager

Page 2 of 2