



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/81019

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer Sudhir Equipment Rental Company

Location Dammam, Kingdom of Saudi Arabia

Description Pressure Relief Valve

Manufacturer HARSHAD

Model No. N/A

Serial No. 2404934

Receipt Date : 16-Feb-26

ID. No. SAC-825-227

Calibration Date : 17-Feb-26

Location Calibration Lab (Dammam)

Calibration Due Date : 16-Aug-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model / Cert. No.	Serial No.	Due Date
High Pressure Pneumatic Pressure Source	SA178189-224312	214E2408011	02-Dec-26
Digital Pressure Calibrator	SA178189-224313	2232411017	02-Dec-26

Calibrated By

64
CALENGR
Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

Customer **Sudhir Equipment Rental Company**
Address Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**

Manufacturer HARSHAD

Model N/A

Serial No. 2404934

ID. No. SAC-825-227

Calibration Date 17-Feb-26

Calibration Due Date 16-Aug-26

Temperature Range 24.2°C

Relative Humidity RH 49%

Test Results

Set Pressure	:	12.00	BAR	Spring Range	:	N/A	
Back Pressure	:	0.00	BAR	Size Inlet/Outlet	:	2"X1	
CDTP	:	12.00	BAR	Type of Valve	:	Metal Seated	
Setting Accuracy	:	± 3 % of Setting			Test Medium	:	Air
Pop Test : PASSED							

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	NO
The instrument was adjusted	YES
The instrument was repaired	No

++++ End of the Certificate +++++

Calibrated By: 
Calibration Engineer

UIC-QP-7.8-24 F02, REV 0

Approved by: 
QA/QC Engineer