



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/DAM/70418

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company		
Address	Dammam, Kingdom of Saudi Arabia		
Description	Crane Anemometer		
Manufacture	SANY		
Model	N/A		
Serial. No.	72319030504		
Plate/ID No.	3783 AAA /SRC-RT-090-022	Calibration Date	: 18-Nov-25
Location	Dammam	Calibration Due Date	: 17-May-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

Reference Standard Equipment Used

Description	Cert. No.	Serial No	Due Date
Mini Wind Tunnel	2023-02	230503	05-Dec-25

Calibrated By
40
CALENGR
Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦
هاتف ٥١٥ ١٣ ٣٦٧ ٩٦٦

CERTIFICATE NO : UIC/DAM/70418

Page

2 of 2

Customer **Sudhir Equipment Rental Company**

Address Dammam, Kingdom of Saudi Arabia

Description **Crane Anemometer**

Calibration Date 18-Nov-25

Manufacturer SANY

Calibration Due Date 17-May-26

Model N/A

Temperature 43.6°C

Serial. No. 72319030504

Relative Humidity RH 36%

Plate/ID No. 3783 AAA /SRC-RT-090-C

Calibration Test Data

Function Tested	Setting Range in Master	UUT Readings	Error of Indication	% Error (FS)
<u>Air Velocity</u> (m/s)	(m/s)	(m/s)	(m/s)	(m/s)
	0.00	0.00	0.000	0.000
	2.50	2.40	-0.100	-0.250
	5.00	4.90	-0.100	-0.250
0	10.00	9.90	-0.100	-0.250
to	15.00	14.90	-0.100	-0.250
	10.00	9.80	-0.200	-0.500
40	5.00	4.90	-0.100	-0.250
	2.50	2.40	-0.100	-0.250
	0.00	0.00	0.000	0.000

General Information about the Calibration

The Crane Anemometer has been Calibrated on the function of Air.

Physical Condition of the UUT

The Unit Under Test is in Good Physical Condition.

Stability

The readings given are the results at the time of calibration and do not carry any implication regarding the Long-term Stability of the instrument.



*** End of the Certificate ***

Calibrated By: _____
Calibration Engineer

UIC-QP-7.8-24 F02, REV 0

Approved by: _____
QA/QC Engineer