

CALIBRATION CERTIFICATE

DETAILS OF CALIBRATION CERTIFICATE

Certificate No	GTEC/AJ/26/NS25112-03	Date of Issue	05.10.2026
Date of Receipt	05.10.2026		
Calibration Date	05.10.2026		
Recommended Calibration Due	04.04.2027		

DETAILS OF CUSTOMER

Customer Name	M/s. Sudhir Equipment Rental Company.
Customer Address	Dammam, Kingdom of Saudi Arabia.

Glory Technology Calibration & Inspection Company (GTEC) certifies that the below described instrument has been calibrated and the measurements are traceable through an unbroken chain to National/International Standards and International System of Units (SI). This calibration certificate documents the traceability of the reference standards used and are available upon the request of customer. The calibration results are documented on the succeeding page/pages of this certificate. The customer is obliged to have the instruments calibrated at regular intervals.

DETAILS OF UNIT UNDER CALIBRATION (UUC)

Description	Pressure Relief Valve
Spring Range / Set Pressure	7.7 bar / 7.7 bar
Make / Model	AFS / A28X-161
Serial No / Identification No	240056
Physical Condition	Satisfactory

DETAILS OF REFERENCE EQUIPMENTS/STANDARDS

Description	Digital Pressure Gauge
Model / Serial No	CPG 1500 / 1A02N5D8M2C
Certificate No / Validity	MTE-CC-SF66000451-4 / 22.10.2026
Traceability	MTE / IAS CL-248

DETAILS OF ENVIRONMENTAL CONDITIONS

Temperature	23.5°C	Relative Humidity	43.2%RH
-------------	--------	-------------------	---------

DETAILS OF CALIBRATION PROCEDURE

Reference Procedure / Standard	GTEC/WI/013 / API 527
--------------------------------	-----------------------

Calibrated By



Goutham J

Calibration Engineer



Authorized By

C. Balasubramanian

Technical Manager



Certificate No :	GTEC/AJ/26/NS25112-03
------------------	-----------------------

CALIBRATION RESULTS

Test Result

Set Pressure	7.7 bar	Test Medium	Water
CDTP	7.7 bar	Valve Inlet Size	3/4"
Start to Leak Pressure	7.7 bar	Valve Outlet Size	--
Seat Leak Test Pressure	6.9 bar	Pressure Setting Tolerance	± 3 %

Physical Inspection Result

Sl. No	Description	Condition	Remarks
1	Valve Body	Good	-
2	Bonnet	Good	-
3	Cap	Good	-
4	Stem	Good	-
5	Adjusting Screw	Good	-
6	Seat	Good	-

Note.

1. The UUC was calibrated at Lab.
2. Calibration Certificate shall not be reproduced, expect in full, without written approval of GTEC.
3. Results reported are valid at the time and under stated environmental conditions of calibration.
4. This Certificate is relates only to the particular item calibrated.

*****End of Certificate*****

Calibrated By



Goutham J

Calibration Engineer



www.gteccalibration.com

Authorized By

C. Balasubramanian

C.Balasubramanian

Technical Manager

Page 2 of 2