



الشركة العالمية لخدمات التفتيش المحدودة

UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س. ت ١١٦٤٣ / ٢٠٥٥

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

## CERTIFICATE OF CALIBRATION

**CERTIFICATE NO: UIC/DAM/109765**

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

**Customer Sudhir Equipment Rental Company**

**Location** Dammam, Kingdom of Saudi Arabia.

**Description Pressure Gauge**

**Manufacturer** Veethree

**Model** N/A

**Serial No.** N/A

**ID. No.** SAC-825-234-PG

**Location** Calibration Lab (Dammam)

**Receipt Date** : 12-Sep-26

**Calibration Date** : 12-Sep-26

**Calibration Due Date** : 11-Mar-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

### Reference Standard Equipment Used

| Description                             | Model / Cert. No. | Serial No   | Due Date  |
|-----------------------------------------|-------------------|-------------|-----------|
| High Pressure Pneumatic Pressure Source | SA178189-224312   | 214E2408011 | 02-Dec-26 |
| Digital Pressure Calibrator             | SA178189-224313   | 2232411017  | 02-Dec-26 |

Calibrated By

Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement results are not to be reproduced except in full without the prior permission of UIC-QP-7.8-24 F02, REV 0

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٨٨٨٢ ١٣ ٣٤٤ / فاكس ٨٨٨١ ١٣ ٣٤٤ +٩٦٦

هاتف ٠٥١٥ ٣٦٧ ١٣ +٩٦٦

CERTIFICATE NO: UIC/DAM/109765

Page

2 of 2

Customer **Sudhir Equipment Rental Company**

Location Dammam, Kingdom of Saudi Arabia.

Description **Pressure Gauge**

Manufacturer Veethree

Model N/A

Serial No. N/A

ID. No. SAC-825-234-PG

Calibration Date 12-Sep-26

Calibration Due Date 11-Mar-27

Temperature Range 23.5°C

Relative Humidity RH 46%

Calibration Test Results For Pressure Gauge

Min. Range: 0.0 BAR

Max. Range: 25.0 BAR

| Function Tested |                          | Gauge Pressure |                            |
|-----------------|--------------------------|----------------|----------------------------|
| UUT Setting     | Actual Reading in Master | Error          | Error (FS)<br>(% of Range) |
| BAR             | BAR                      | BAR            |                            |
| 0.00            | 0.00                     | 0.000          | 0.000                      |
| 5.00            | 4.97                     | -0.030         | -0.120                     |
| 10.00           | 9.95                     | -0.050         | -0.200                     |
| 15.00           | 14.95                    | -0.050         | -0.200                     |
| 20.00           | 19.93                    | -0.070         | -0.280                     |
| 25.00           | 24.95                    | -0.050         | -0.200                     |

Calibration Procedure:

UIC/P/PPGCP Rev. 00

Reference Standard:

As Generally given in MSL Technical guide & Tolerance as per EN 837 standard & based on the Manufacture's Specifications.

Remarks / Deviations IF ANY:

The results stated in this certificate relate only to the item calibrated.  
The item was received in good condition.



++++ End of the Certificate +++++



Calibrated By: \_\_\_\_\_  
Calibration Engineer

Approved by:   
QA/QC Engineer

UIC-QP-7.8-24 F02, REV 0