



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No.2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/111575

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company		
Location	Dammam, Kingdom of Saudi Arabia		
Description	Pressure Relief Valve		
Manufacturer	AFS Safety Valve		
Type	AK28X-16T		
Serial No.	260088	Receipt Date	28-Sep-26
ID. No.	SAC-850-504	Calibration Date	28-Sep-26
Location	Calibration Lab (Dammam)	Calibration Due Date	27-Mar-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

Reference Standard Equipment Used


Description	Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA178189-224312	214E2408011	02-Dec-26
Digital Pressure Calibrator	SA178189-224313	2232411017	02-Dec-26

Calibrated By

Calibration Engineer



Approved By


QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

UIC-OP-7.8-24-F02-REV 0

This certificate and the attached measurement result sheets must not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦
هاتف ٥١٥ ١٣ ٣٦٧ ٩٦٦

CERTIFICATE NO: UIC/DAM/111575

Page

2 of 2

Customer **Sudhir Equipment Rental Company**
Location Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**

Manufacturer AFS Safety Valve

Model AK28X-16T

Serial No. 260088

ID. No. SAC-850-504

Calibration Date 28-Sep-26

Calibration Due Date 27-Mar-27

Temperature Range 23.5°C

Relative Humidity RH 47%

Test Results

Set Pressure : 12.00 BAR

Maximum Presssure : N/A

Back Pressure : 0.00 BAR

Size Inlet/Outlet : 1" x 1½"

CDTP : 12.00 BAR

Setting Accuracy : ± 3 % of Setting

Type of Valve : Metal Seated

Pop Test : PASSED

Test Medium : AIR

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)

No

The instrument was adjusted

Yes

The instrument was repaired

No



++++ End of the Certificate +++++



Calibrated By:

Calibration Engineer

UIC-QP-7.8-24 F02, REV 0

Approved by:

QA/QC Engineer