الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.
CR No. 2055011643
س. ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/RIY/48689

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	**Sudhir Equipment Rental Company.**			
Address	RIYADH, Kingdom of Saudi Arabia			
Description	**Pressure Relief Valve**			
Manufacturer	HARSHAD	**Receipt Date**	:	10-Oct-26
Part No.	0832100066	**Calibration Date**	:	10-Oct-26
Serial No.	2424233	**Report Date**	:	10-Oct-26
ID No.	SAC-500-310	**Issue date**	:	10-Oct-26
Location	Calibration Lab	**Calibrtation Due Date**	:	9-Apr-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Model/Cert No	Serial No	Due Date
High Pressure Pneumatic Calibrator	SA178189-223867	2011H13013	23-Nov-26
Digital Pressure Calibrtor	SA178189-223866	223320200723	26-Nov-26

Calibrated By

Calibration Engineer

UIC-QP-7.8.2-17-REV 02

Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515
www.ui.com.sa
ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ٣٤٤ ١٣ ٩٦٦+ / فاكس ٨٨٨١ ٣٤٤ ١٣ ٩٦٦+
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦+

CERTIFICATE NO : UIC/RIY/48689

Customer	**Sudhir Equipment Rental Company.**		
Address	RIYADH, Kingdom of Saudi Arabia		
Description	**Pressure Relief Valve**	**Calibration Date**	10-Oct-26
Manufacturer	HARSHAD	**Calibration Due Date**	9-Apr-27
Model	0832100066	**Temperature**	23.0 ± 2°C
Serial No.	2424233	**Relative Humidity**	RH 45% to 75%
ID No.	SAC-500-310		

Test Results

Set Pressure :	**10.0**	**BAR**	**Spring Range**	**: N/A**
Back Pressure:	**0.0**	**BAR**	**Size Inlet / Outlet**	**: 3/4"X 1"**
CDTP :	**10.0**	**BAR**		
Setting Accuracy : ± 3 % of Setting			**Type of Valve**	**: Metal Seated**
Pop Test : PASSED			**Test Medium**	**: Air**

BASIS OF TEST:

Ref Standard ASME Sec VIII,Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic / pnenumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company Maintains traceabilty to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	No
The instrument was adjusted	Yes
The instrument was repaired	No

Result:

Calibration Result is pass.

✧✧✧✧✧ End of the Certificate ✧✧✧✧✧

Calibrated By: _____
Calibration Engineer

Approved by: _____
QA/QC Engineer