



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC / TUR / 25255

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Sudhir Equipment Rental Company**
Address DAMMAM, KINGDOM OF SAUDI ARABIA.
Description **Pressure Relief Valve**
Manufacturer ATLAS COPCO INDIA LTD
Model/Type 8321000066
Serial No. 2307812
ID. No. SAC-825-186 **Calibration Date** : 25-Apr-26
Location Calibration Lab (Turaif) **Calibration Due Date** : 24-Oct-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Cert. No./Model	Serial No	Due Date
PORTABLE HIGH PRESSURE PNEUMATIC HAND PUMP	VIC-25-42537	2011H13013	18-Sep-26
DIGITAL PRESSURE GAUGE	VIC-25-57999	S132001185	18-Sep-26

Calibrated By

3
CALENGR

Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦
هاتف ٠٥١٥ ١٣ ٩٦٦

CERTIFICATE NO: UIC / TUR / 25255

Customer	Sudhir Equipment Rental Company		
Address	DAMMAM, KINGDOM OF SAUDI ARABIA.		
Description	Pressure Relief Valve	Calibration Date	25-Apr-26
Manufacturer	ATLAS COPCO INDIA LTD	Calibration Due Date	24-Oct-26
Model/Type	8321000066	Lab Temperature	23°C
Serial No.	2307812	Lan Relative Humidity	45%
ID. No.	SAC-825-186		

Test Results

Set Pressure :	11 BARG	Spring Range : -----
Back Pressure :	0 BARG	Size Inlet/Outlet : 1"ISO7RM x 1.5"BSPF
CDTP :	11 BARG	Type of Valve : Metal Seated
Setting Accuracy : $\pm 3\%$ of Setting		Test Medium : Air
Pop Test : PASSED		

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed below for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	Yes
The instrument was adjusted	Yes
The instrument was repaired	No

Calibrated By:



Calibration Engineer



Approved by:

QA/QC Engineer