



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO : UIC/JED/79906

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **SUDHIR EQUIPMENT RENTAL COMPANY.**
Address JEDDAH - KINGDOM OF SAUDI ARABIA

Description **SAFETY RELIEF VALVE**

Manufacturer HARSHAD
Part No. 9095799400


Serial No. 2408692 Receipt Date : 30-Apr-26
ID No. SAC-825-245 Calibration Date : 30-Apr-26
Location Calibration Lab (Jeddah) Calibration Due Date : 29-Oct-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Certificate No.	Serial No.	Due Date
High Pressure Pneumatic Calibrator	SA19-9047-75821	214E2408004	24-Dec-26
Digital Pressure Gauge	SA19-9047-75820	2232406117	24-Dec-26

Calibrated By


Calibration Engineer



UIC-QP-7.8-24 F02, REV 0

Approved By


QA/QC Engineer

PAGE : 1 of 2

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٣٤٤ / فاكس ٨٨٨١ ١٣ ٣٤٤ +٩٦٦
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦ +

CERTIFICATE NO : UIC/JED/79906

PAGE

2 of 2

Customer Address	SUDHIR EQUIPMENT RENTAL COMPANY. JEDDAH - KINGDOM OF SAUDI ARABIA		
Description	SAFETY RELIEF VALVE		
Manufacturer	HARSHAD	Calibration Date	30-Apr-26
Part No.	9095799400	Calibration Due Date	29-Oct-26
Serial No.	2408692	Temperature	21.6°C
ID No.	SAC-825-245	Relative Humidity	47%

Test Results

Set Pressure :	9.5 Bar	Spring No./ Press. Range : ----
Back Pressure:	0.0 Bar	Size Inlet : 3/4"
CDTP :	9.5 Bar	Type of Valve : Metal Seated
Setting Accuracy : $\pm 3\%$ of Setting		Test Medium : Air/Gas/Water/Oil
Pop Test : PASSED		

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Safety relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed below for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	No
The instrument was adjusted	Yes
The instrument was repaired	No

++++ End of the Certificate +++++



Calibrated By: _____
Calibration Engineer

UIC-QP-7.8-24 F02, REV 0

Approved by: _____
QA/QC Engineer