الشركة العالمية لخدمات التفتيش المحدودة
UNIVERSAL INSPECTION CO. LTD.
CR No.2055011643 س.ت ٢٠٥٥٠١١٦٤٣

# AIR QUALITY TEST REPORT

**REPORT NO: UIC/AIR/2026/2088**

**DATE OF REPORT: 01 NOV 2026**
**DATE OF TESTING: 01 NOV 2026**

## INTRODUCTION:

A cylinder containing breathing air was received from **M/S SUDHIR EQUIPMENT RENTAL COMPANY,** on 31st OCT 2026. It was requested by the client to carryout purity testing as per CGA Standard Grade "D" breathing air.

| | | |
|---|---|---|
| **Customer Name** | : | **SUDHIR EQUIPMENT RENTAL COMPANY,** |
| **Location** | : | **DAMMAM, KSA** |
| **Sample Identification** | : | **AIR COMPRESSOR** |
| **Sticker Number** | : | **AQT/2390** |
| **Make** | : | **ATLAS COPCO** |
| **Model** | : | **XAS500-175** |
| **SL.No** | : | **PE2560326** |
| **ID.No** | : | **SAC-500-350** |

## TEST PROCEDURE:

The major components of the air sample were determined by using gas chromatograph equipped with thermal conductivity and flame ionization detectors. The other impurities are analyzed by Chemical reaction method using detection tubes.

## SPECIFICATIONS:

CGA Commodity Specification for Air, ANSI/CGA G-7.1-2018

**RESULTS**

| Test Description | Unit | Results | CGA Grade - D Limits (Max) |
|---|---|---|---|
| Oxygen ( $O_2$ ) | (Vol. %) | 22.35 | 19.5 – 23.5 |
| Carbon Dioxide ( $CO_2$) | (Vol. %) | 0.04 | 0.10 |
| Carbon Monoxide (CO) | (ppm) | < 2.0 | 10.0 |
| Oil Condensed | ($mg/m^3$) | < 1.0 | < 5.0 |
| Moisture Content | (ppm) | 59 | 67.0 |
| Odour | --- | None | None |

## REMARKS:

The above test result of the air sample complies with the Compressed Gas Association Standard for Grade "D" Breathing Air for SCBA or Airline respirators. Further precautionary measure time to time, we would recommend to clean the filter tap prior to use in breathing purposes.

**Tested By:**

**Materials Analysis Dept.**

**Approved By:**

**QA/QC Manager**



This report relates only to the sample tested and only shall be reproduced in full with written approval of UIC testing laboratory

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515
www.ui.com.sa
ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ٣٤٤ ١٣ ٩٦٦+ / فاكس ٨٨٨١ ٣٤٤ ١٣ ٩٦٦+
هاتف ٠٥١٥ ٣٦٧ ١٣ ٩٦٦+