



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC / 459230

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company		
Address	Al - Jubail, Kingdom of Saudi Arabia.		
Description	Pressure Relief Valve		
Manufacturer	HARSHAD		
Part/Model	9095799400		
Serial No.	2404853	Receipt Date	: 08-Jul-26
ID No	N/A	Calibration Date	: 09-Jul-26
Location	Calibration Lab (Jubail)	Calibration Due Date	: 08-Jan-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements


Reference Standard Equipment Used

Description	Cert. No./Model	Serial No	Due Date
High Pressure Pneumatic Comparator	SA178189-224615	214E2207005	14-Jan-27
Digital Pressure Calibrator	30500	2232209008	20-Apr-27

Calibrated By

Calibration Engineer



Approved By

Quality / Lab Manager

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement results are not to be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١
هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

Customer **Sudhir Equipment Rental Company**
Address Al - Jubail, Kingdom of Saudi Arabia.

Description	Pressure Relief Valve	Calibration Date	09-Jul-26
Manufacturer	HARSHAD	Calibration Due Date	08-Jan-27
Part/Model	9.10E+09	Temperature	24.5 °C
Serial No.	2404853	Relative Humidity	RH 48 %

Test Results

Set Pressure :	9.5	BARG	Spring Range	: N/A
Back Pressure:	0.0	BARG	Size Inlet/outlet	: 3/4"
COTP :	9.5	BARG	Type of Valve	: Metal Seated
Setting Accuracy : ± 3 % of Setting			Test Medium	: Air
Pop Test : PASSED				

BASIS OF TEST:

Ref Standard API 527, and ASME Sec VIII, Div 1.

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and using a reference standard pressure gauge or test gauge and a hydrostatic /pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company Maintains traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	No
The instrument was adjusted	No
The instrument was repaired	No



◆◆◆◆◆ End of the Certificate ◆◆◆◆◆

Calibrated By:



Calibration Engineer

UIC-QP-7.8 24 F02, REV 0

Approved by:

Quality / Lab Manager