



الشركة العالمية لخدمات التفتيش المحدودة

UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC / SAF / 109108

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company		
Address	Dammam, Kingdom of Saudi Arabia.		
Description	Pressure Relief Valve		
Manufacturer	HARSHAD		
Part No.	9095799400		
Serial No.	2413103	Receipt Date	: 13-Jul-26
ID No.	SAC-250-279	Calibration Date	: 13-Jul-26
Location	Calibration Lab (Safaniya)	Calibration Due Date	: 12-Jan-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used

Description	Cert. No.	Serial No	Due Date
High Pressure Pneumatic Comparator	28014	2000E2208001	06-Apr-27
Digital Pressure Gauge	28013	2232209002	06-Apr-27

Calibrated By

Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٣٤٤ / فاكس ٨٨٨١ ١٣ ٣٤٤ +٩٦٦
هاتف ٥١٥ ١٣ ٣٦٧ +٩٦٦

CERTIFICATE NO: UIC / SAF / 109108		PAGE	2 of 2
Customer	Sudhir Equipment Rental Company		
Address	Dammam, Kingdom of Saudi Arabia.		
Description	Pressure Relief Valve		
Manufacturer	HARSHAD	Calibration Date	13-Jul-26
Part No.	9095799400	Calibration Due Date	12-Jan-27
Serial No.	2413103	Temperature	23.0 ± 2°C
ID No.	SAC-250-279	Relative Humidity	RH 45% to 75%
<u>Test Results</u>			
Set Pressure :	9.5 BARG	Spring Range	: N/A
Back Pressure:	0.0 BARG	Size Inlet	: 3/4"
CDTP :	9.5 BARG	Type of Valve	: Metal Seated
Setting Accuracy :	± 3 % of Setting		
Pop Test :	PASSED		
<u>BASIS OF TEST:</u>			
Ref Standard ASME Sec VIII, Div I			
<u>Summary of Test Procedure:</u>			
The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.			
<u>Traceability:</u>			
The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.			
<u>Condition of Instrument</u>			
With in the tolerance on receipt(as found)	No		
The instrument was adjusted	No		
The instrument was repaired	No		
 Calibrated By: _____ Calibration Engineer ++++ End of the Certificate +++++  UIC-QP-7-8-24 F02, REV 0 Approved by: _____ QA/QC Engineer			