



الشركة العالمية لخدمات التفتيش المحدودة

UNIVERSAL INSPECTION CO. LTD.

CR No. 2053011643

س.ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC / 448435

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer **Sudhir Equipment Rental Company**

Address Al Jubail, Kingdom of Saudi Arabia.

Description **Pressure Relief Valve**

Manufacturer EW

Model/Type N/A

Serial No. SAC-500-261

ID No. N/A

Location Calibration Lab (Jubail)

Receipt Date : 30-May-26

Calibration Date : 31-May-26

Calibration Due Date : 30-Nov-26

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO 17025 requirements

Reference Standard Equipment Used


Description	Model / Cert. No.	Serial No	Due Date
High Pressure Pneumatic Comparator	AME/25/TC-0711-A1	214E2408017	10-Jul-26
Digital Pressure Calibrator	AME/25/CC-0703-A2	2232209008	02-Jul-26

Calibrated By



Calibration Engineer

Approved By



Quality / Lab Manager

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement result sheets shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A. UIC-QP-7-8-24 F02, REV 0

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO: UIC / 448435

Customer
Address**Sudhir Equipment Rental Company**
Al Jubail, Kingdom of Saudi Arabia.Description
Manufacturer
Model/Type
Serial No.
ID No.**Pressure Relief Valve**
EW
N/A
SAC-500-261
N/ACalibration Date
Calibration Due Date
Temperature
Relative Humidity
31-May-26
30-Nov-26
24.8 °C
RH 49 %Test Results

Set Pressure : 10.00 BARG

Spring Range : N/A

Back Pressure: 0.00 BARG

Size Inlet/Outlet : 1/2"X1"

CDTP : 10.00 BARG

Type of Valve : Metal Seated

Setting Accuracy : ± 3 % of Setting

Test Medium : Air

Pop Test : PASSED

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and using a reference standard pressure gauge or test gauge and a hydrostatic /pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company Maintains traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)

No

The instrument was adjusted

No

The instrument was repaired

No

***** End of the Certificate *****



Calibrated By:

Calibration Engineer



UIC-QP-7.8-24 F02, REV 0

Approved by:

Quality / Lab Manager