



الشركة العالمية لخدمات التفتيش المحدودة

UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س. ت ٢٠٥٥٠١١٦٤٣

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/106499

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer Sudhir Equipment Rental Company

Location Dammam, Kingdom of Saudi Arabia

Description Pressure Relief Valve

Manufacturer AFS Safety Valve

Type AK28X-16T

Serial No. 260026

Receipt Date : 19-Aug-26

ID. No. SAC-850-489

Calibration Date : 19-Aug-26

Location Calibration Lab (Dammam)

Calibration Due Date : 18-Feb-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

Reference Standard Equipment Used

Description	Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA178189-224312	214E2408011	02-Dec-26
Digital Pressure Calibrator	SA178189-224313	2232411017	02-Dec-26

Calibrated By

Calibration Engineer

Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement results shall not be reproduced except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.

Tel.: +966 13 344 8882 / Fax: +966 13 344 8881

Tel.: +966 13 367 0515

www.ui.com.sa

ص.ب. ٢٠٧٢، الجبيل ٣١٩٥١ - المملكة العربية السعودية

هاتف ٩٦٦ ١٣ ٣٤٤ ٨٨٨٢ / فاكس ٩٦٦ ١٣ ٣٤٤ ٨٨٨١

هاتف ٩٦٦ ١٣ ٣٦٧ ٠٥١٥

CERTIFICATE NO: UIC/DAM/106499

Page

2 of 2

Customer **Sudhir Equipment Rental Company**
Location Dammam, Kingdom of Saudi Arabia

Description **Pressure Relief Valve**
Manufacturer AFS Safety Valve
Model AK28X-16T
Serial No. 260026
ID. No. SAC-850-489

Calibration Date 19-Aug-26
Calibration Due Date 18-Feb-27
Temperature Range 23.5°C
Relative Humidity RH 47%

Test Results

Set Pressure	:	12.00	BAR	Maximum Pressure	: N/A
Back Pressure	:	0.00	BAR	Size Inlet/Outlet	: 1" X 1½"
CDTP	:	12.00	BAR	Type of Valve	: Metal Seated
Setting Accuracy	:	± 3 % of Setting		Test Medium	: AIR
Pop Test : PASSED					

BASIS OF TEST:

Ref Standard ASME Sec VIII, Div I

Summary of Test Procedure:

The Calibration of above Pressure relief valve is performed through the verification of set pressure and seat tightness using a reference standard pressure gauge or test gauge and a hydrostatic / pneumatic test system.

Traceability:

The reported tests were undertaken using equipment detailed mentioned in Reference Standard Equipment's for which the company maintains Traceability to the third party.

Condition of Instrument

With in the tolerance on receipt(as found)	No
The instrument was adjusted	Yes
The instrument was repaired	No



++++ End of the Certificate +++++



Calibrated By:

Calibration Engineer

UIC-QP-7.8-24 F02, REV 0

Approved by:

QA/QC Engineer