



الشركة العالمية لخدمات التفتيش المحدودة UNIVERSAL INSPECTION CO. LTD.

CR No. 2055011643

س. ت ١١٦٤٣ ٢٠٥٥٠

INSTRUMENT & EQUIPMENT CALIBRATION DIVISION

CERTIFICATE OF CALIBRATION

CERTIFICATE NO: UIC/DAM/107312

This certifies that the instrument below has been tested/calibrated using equipment traceable to National/International Standards or verified by ratio type measurements using self calibrating techniques.

Information on the calibration process and the results are produced below and given in the accompanying sheets of this document.

All related traceable documents are on file at **UNIVERSAL INSPECTION CO. LTD.** and may be examined on request

Customer	Sudhir Equipment Rental Company		
Location	Dammam, Kingdom of Saudi Arabia		
Description	Pressure Gauge		
Manufacturer	VEETHREE		
Model	N/A		
Serial No.	N/A	Receipt Date	: 26-Aug-26
ID. No.	SAC-500-179	Calibration Date	: 26-Aug-26
Location	Calibration Lab (Dammam)	Calibration Due Date	: 25-Feb-27

The instrument was calibrated using the following reference, maintained at UIC Instrument & Equipment Calibration division conforming to ISO/IEC 17025:2017 requirements

Reference Standard Equipment Used

Description	Cert. No.	Serial No	Due Date
High Pressure Pneumatic Pressure Source	SA178189-224312	214E2408011	02-Dec-26
Digital Pressure Calibrator	SA178189-224313	2232411017	02-Dec-26

Calibrated By

Calibration Engineer



Approved By

QA/QC Engineer

The measurement results pertain only to the instrument calibrated.

This certificate and the attached measurement results are to be used except in full without the prior permission

P.O. Box No. 2072, Jubail - 31951, K.S.A.
Tel.: +966 13 344 8882 / Fax: +966 13 344 8881
Tel.: +966 13 367 0515

www.ui.com.sa

ص. ب ٢٠٧٢ ، الجبيل ٣١٩٥١ - المملكة العربية السعودية
هاتف ٨٨٨٢ ١٣ ٩٦٦ / فاكس ٨٨٨١ ١٣ ٩٦٦
هاتف ٠٥١٥ ٣٦٧ ٩٦٦

CERTIFICATE NO: UIC/DAM/107312

Page

2 of 2

Customer **Sudhir Equipment Rental Company**

Location Dammam, Kingdom of Saudi Arabia

Description **Pressure Gauge**

Manufacturer VEETHREE

Model N/A

Serial No. N/A

ID. No. SAC-500-179

Calibration Date 26-Aug-26

Calibration Due Date 25-Feb-27

Temperature Range 23.5°C

Relative Humidity RH 48%

Calibration Test Results for Pressure Gauge

Min. Range: 0.0 BAR

Max. Range: 25.0 BAR

Function Tested		Gauge Pressure	
UUT Setting	Actual Reading in Master	Error	Error (FS)
BAR	BAR	BAR	(% of Range)
0.00	0.00	0.000	0.000
5.00	5.02	0.020	0.080
10.00	10.03	0.030	0.120
15.00	15.04	0.040	0.160
20.00	20.02	0.020	0.080
25.00	25.03	0.030	0.120

Calibration Procedure:

UIC/P/PPGCP Rev. 00

Reference Standard:

As Generally given in MSL Technical guide & Tolerance as per EN 837 standard & based on the Manufacture's Specifications.

Remarks / Deviations IF ANY:

The results stated in this certificate relate only to the item calibrated.
The item was received in good condition.

++++ End of the Certificate +++++



Calibrated By: _____

Calibration Engineer

UIC-QP-7.8-24 F02, REV 0

Approved by: _____

QA/QC Engineer