

Ordering information

1 Volumetric flasks

Volumetric flasks, BLAUBRAND®, class A, DE-M

Boro 3.3. DIN EN ISO 1042. Calibrated to contain (TC, In).

Capacity ml	Description	Closure option	Cat. No.
5 - 10000	with ISO batch certificate	PP/PE stopper	37233 - 37293
	with ISO individual certificate	PP/PE stopper	937233 - 937293
5 - 5000	with ISO batch certificate	glass stopper	37256 - 37294
	with ISO individual certificate	glass stopper	937256 - 937294
10 - 1000	with ISO batch certificate, beaded rim	without stopper	37045 - 37053
	with ISO individual certificate, beaded rim	without stopper	937045 - 937053
5 - 2000	with USP batch certificate	PP stopper	36938 - 36954
	with USP individual certificate	PP stopper	956938 - 956954
	with USP batch certificate	glass stopper	36968 - 36984
	with USP individual certificate	glass stopper	956968 - 956984



Volumetric flasks, BLAUBRAND®, class A, DE-M, amber

Boro 3.3. DIN EN ISO 1042. Calibrated to contain (TC, In).

Capacity ml	Description	Closure option	Cat. No.
5 - 1000	with ISO batch certificate	PP stopper	37401 - 37413
	with ISO individual certificate	PP stopper	937401 - 937413
	with ISO batch certificate	glass stopper	37438 - 37453
	with ISO individual certificate	glass stopper	937438 - 937453
5 - 1000	with USP batch certificate	PP stopper	37481 - 37491
	with USP individual certificate	PP stopper	957481 - 957491
	with USP batch certificate	glass stopper	37461 - 37471
	with USP individual certificate	glass stopper	957461 - 957471





Volumetric flasks, BLAUBRAND® ETERNA, class A, DE-M

Boro 3.3. To DIN EN ISO 1042. Calibrated to contain (TC, In).

Capacity ml	Description	Closure option	Cat. No.
5 - 2000	with ISO batch certificate	PP stopper	36838 - 36854
	with ISO individual certificate	PP stopper	936838 - 936854
	with USP individual certificate	PP stopper	956841 - 956854



Volumetric flasks, BLAUBRAND®, class A, DE-M, PURprotect

Boro 3.3. DIN EN ISO 1042. Calibrated to contain (TC, In).

Capacity ml	Description	Closure option	Cat. No.
50 - 1000	with ISO batch certificate	PP stopper	36548 - 36553



Volumetric flasks, BLAUBRAND®, class A, DE-M, trapezoidal


Boro 3.3. DIN EN ISO 1042. Calibrated to contain (TC, In).

Capacity ml	Description	Closure option	Cat. No.
1 - 50	with ISO batch certificate	PP stopper	36401 - 36408
	with ISO individual certificate	PP stopper	936401 - 936408

2 Graduated cylinders

Graduated cylinders, tall form, BLAUBRAND®, class A, DE-M

Boro 3.3. To DIN EN ISO 4788. Calibrated to contain (TC, In). With spout and hexagonal glass base.



Capacity ml	Description	Cat. No.
5 - 2000	with ISO batch certificate	32105 - 32164
	with USP batch certificate	32805 - 32864

Graduated cylinders, tall form, BLAUBRAND® ETERNA, class A, DE-M

Boro 3.3. To DIN EN ISO 4788. Calibrated to contain (TC, In). With spout and hexagonal glass base.

Capacity ml	Description	Cat. No.
5 - 2000	with ISO batch certificate	32705 - 32764

**Mixing cylinders, tall form, class A, DE-M**

Boro 3.3. DIN EN ISO 4788. Calibrated to contain (TC, In).

Capacity ml	Description	Cat. No.
10 - 1000	with ISO batch certificate	32408 - 32462



3 Bulb pipettes

Bulb pipettes, BLAUBRAND®, class AS, DE-M

AR-GLAS®. DIN EN ISO 648. Calibrated to deliver (TD, Ex). Waiting time 5 seconds.

Capacity ml	Description	Cat. No.
0,5 - 100	1 mark, with ISO batch certificate	29701 - 29719
	1 mark, with USP batch certificate	30601 - 30619
0,5 - 50	2 marks, with ISO batch certificate	29721 - 29738


**Bulb pipettes, BLAUBRAND® ETERNA, class AS, DE-M**

AR-GLAS®. DIN EN ISO 648. Calibrated to deliver (TD, Ex). Waiting time 5 seconds.

Capacity ml	Description	Cat. No.
1 - 50	with ISO batch certificate	30502 - 30518



4 Graduated pipettes



Graduated pipettes, BLAUBRAND®, DE-M

AR-GLAS®. DIN EN ISO 835. Waiting time 5 seconds.

Capacity ml	Description	Cat. No.
0,5 - 25	Type 1, class AS, with ISO batch certificate	27721 - 27731
0,5 - 50	Type 2, class AS, with ISO batch certificate	27816 - 27829
	Type 3, class AS, with ISO batch certificate	27705 - 27716
1 - 10	Type 2, class AS, with USP batch certificate	27506 - 27513
0,1 - 0,2	Type Einguss, class A, with ISO batch certificate	27702 - 27704



Graduated pipettes, BLAUBRAND® ETERNA, class AS, DE-M

AR-GLAS®. DIN EN ISO 835. Calibrated to deliver (TD, Ex). Waiting time 5 seconds.

Capacity ml	Description	Cat. No.
0,5 - 25	Type 3, with ISO batch certificate	27316/27319 28405 - 28415

5 Burettes



Burettes, BLAUBRAND®, class AS, DE-M

Boro 3.3. To DIN EN ISO 385. Calibrated to deliver (TD, Ex), PTFE stopcock.

Capacity ml	Description	Cat. No.
10 - 50	lateral valve cock, blue Schellbach stripe	13844 - 13848
	straight valve cock, blue Schellbach stripe	12484 - 12488



Micro-burettes (Bang), BLAUBRAND®, class AS, DE-M

Boro 3.3. To DIN EN ISO 385. Calibrated to deliver (TD, Ex).

Capacity ml	Description	Cat. No.
2 - 10	lateral valve cock, blue Schellbach stripe, with ISO batch certificate	24596 - 24600
	straight valve cock, blue Schellbach stripe, with ISO batch certificate	24265 - 24269

Compact-Burettes, BLAUBRAND®, class AS, DE-M

Boro 3.3, To DIN EN ISO 385. Calibrated to deliver (TD, Ex).

Capacity ml	Description	Cat. No.
10 - 50	blue Schellbach stripe, with ISO batch certificate	13913 - 13918
	white graduated, amber, with ISO batch certificate	13963 - 13968

**6 Compact-Automatic burettes****Compact-Automatic burettes, BLAUBRAND®, class AS, DE-M**

DIN EN ISO 385. Calibrated to deliver (TD, Ex).

Capacity ml	Description	Cat. No.
10 - 50	blue Schellbach stripe, with ISO batch certificate	23923 - 23925
	with PE-bottle, blue Schellbach stripe, with ISO batch certificate	23819 - 23821
	white graduated, amber, with ISO batch certificate	23939 - 23941
	with PE-bottle, white graduated, amber, with ISO batch certificate	23870 - 23872

**Automatic burettes Pellet pattern, BLAUBRAND®, class AS, DE-M**

DIN EN ISO 385. Boro 3.3, calibrated to deliver (TD, Ex).

Capacity ml	Description	Cat. No.
10 - 50	with intermediate stopcock, blue Schellbach stripe	22761 - 22763
	without intermediate stopcock, blue Schellbach stripe	22521 - 22523

