EEG Signal Quality Report

Automated per-channel quality assessment with PASS/FAIL thresholds

FILE: demo_recording.easy
DEVICE: ENOBIO
CHANNELS: 8 SAMPLING RATE: 500 Hz
DURATION: 10.0 s WINDOW: 0.0 – 10.0 s
DATE: April 09, 2026


8 of 8 channels FAIL (0 pass)

1. Per-Channel Metrics


Channel	Mean (µV)	RMS raw (µV)	RMS notch (µV)	RMS filt (µV)	Line 50/60 (dB)	Kurtosis	Events (/s)	Result
Fp1	2	13.9*	13.9*	12.9*	-8.0	14.9*	80.0*	FAIL
Fp2	2	13.5*	13.4*	12.4*	-7.0	13.3*	81.2*	FAIL
F3	-0	7.5*	7.5*	5.9*	-9.0	2.5	114.5*	FAIL
F4	0	7.6*	7.5*	6.0*	-9.4	2.5	114.5*	FAIL
C3	0	8.7*	8.6*	7.3*	-8.0	2.4	101.1*	FAIL
C4	0	8.7*	8.6*	7.3*	-8.1	2.3	101.0*	FAIL
O1	0	15.2*	15.1*	14.4*	-5.9	1.9	46.7*	FAIL
O2	0	15.1*	15.0*	14.3*	-7.5	1.9	49.2*	FAIL

* = exceeds threshold. RMS in µV. Line power in dB. Event threshold: 5.0 µV.


2. Power Spectral Density — page1



3. Filtered — Page1



4. Raw — Page1


Raw Data - Channels 1-8
Ch 1 (µV)
Ch 2 (µV)
Ch 3 (µV)
Ch 4 (µV)
Ch 5 (µV)
Ch 6 (µV)
Ch 7 (µV)
Ch 8 (µV)
10
0
−10
0
2
4
6
8
10
Time (s)