





Quality Inspection Report (Inprocess Inspection)

Report No	QIRI20250386	Select the Appropriate Inspection Stage / Process
Date	18-01-2025	Inspection of Cut length Wire
Production Order	PRO024-2436	
Item Code	Ø 0.90 CL Wire (RAJARATHNA)	
Product Description	Dia 0.90 mm Cut Length Wire 90 mm (RAJARATHNA)	
DHR No	20250205	
Batch No	WC0006/2025	
Reference Drawing	NA	

Sl.N o.	Parameter	Requirement	Instrument Used	Setup Approval (Actual Readings/Observation)										Result (Complied / Not Complied)	Remarks
				1	2	3	4	5	6	7	8	9	10		
1	Diameter	0.90 ± 0.005 mm	DM	0.904	0.903	0.903	0.903	0.903	0.903	0.904	0.903	0.902	0.904	Complied	2
2	Length of cut Wires	90 ± 0.50 mm	DVC	89.69	89.87	89.99	90.09	89.96	89.57	89.97	89.74	89.92	89.85	Complied	
3	Straightness	Straight	Rolling Method	OK	OK	OK	OK	OK	OK	OK	OK	OK	OK	Complied	
4	Defects Free	Free from rust, burr,Damage.	Microscope	OK	OK	OK	OK	OK	OK	OK	OK	OK	OK	Complied	
Inspected By (Name)		LAKSHMI.					Signature (with date)				18-01-2025				

Quality Inspection Report (Inprocess Inspection)

Sl.N o.	Parameter	Date and Time	Instrument Used	Inprocess Samples Inspection (Put "V" - Complied and "X" - Not Complied)										Inspected by (Signature)	Remarks
				1	2	3	4	5	6	7	8	9	10		
1 2 3 4	Diameter Cutting Length Straightness Defects free	18-01-2025	DMM DVC Rolling Method Microscope	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		N/A
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
		12:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
		01:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
		02:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
		03:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓			
Sample Inspected Quantity		✓	Accepted Batch Quantity		✓	Remarks		✓							
Defects Observed (Type)		✓													
Post Quality Inspection Result				Inspected By (Name)		LAKSHMI.		Approved By (Name)		MURIGESHAPPA R					
<input checked="" type="checkbox"/> Batch conform to the requirement and approved <input type="checkbox"/> Batch do not conform to the requirement and not approved				Signature with date		 18-01-2025		Signature with date		 18-01-2025					
Note: Annexure to be provided for additional test report (if required)															

Quality Inspection Report (Inprocess Inspection)

Sl.N o.	Parameter	Date and Time	Instrument Used	Inprocess Samples Inspection (Put "v" - Complied and "X" - Not Complied)										Inspected by (Signature)	Remarks
				1	2	3	4	5	6	7	8	9	10		
1 2 3 4	Diameter Cutting Length Straightness Defects free	04:00 PM	DMM DVC Rolling Method Microscope	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	<i>[Signature]</i>	NA
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		05:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		06:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		20-01-2025		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		09:40 AM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓					
10:40 AM	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓					
	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓					
	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓					
11:40 AM	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓					
Sample Inspected Quantity		Accepted Batch Quantity				Remarks									
Defects Observed (Type)															
Post Quality Inspection Result				Inspected By (Name)		LAKSHMI.		Approved By (Name)		MURIGESHAPPA R					
<input checked="" type="checkbox"/> Batch conform to the requirement and approved				Signature		<i>[Signature]</i>		Signature		<i>[Signature]</i>					
<input type="checkbox"/> Batch do not conform to the requirement and not approved				with date		20-01-2025		with date		20-01-2025					
Note: Annexure to be provided for additional test report (if required)															

Sl.N o.	Parameter	Date and Time	Instrument Used	Inprocess Samples Inspection (Put "v" - Complied and "X" - Not Complied)										Inspected by (Signature)	Remarks
				1	2	3	4	5	6	7	8	9	10		
1 2 3 4	Diameter Cutting Length Straightness Defects free	12:40 PM	DMM DVC Rolling Method Microscope	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	Raj Raj Raj Raj Raj Raj Raj Raj Raj Raj Raj	NA
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
		01:40 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
		02:40 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
		04:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓			
				✓	✓	✓	✓	✓	✓	✓	✓	✓			
Sample Inspected Quantity		500 PCS.	Accepted Batch Quantity		36200 PCS.		Remarks								
Defects Observed (Type)															
Post Quality Inspection Result				Inspected By (Name)		LAKSHMI.		Approved By (Name)		MURIGESHAPPA R					
<input checked="" type="checkbox"/> Batch conform to the requirement and approved				Signature		Raj 20-01-2024		Signature		Raj 20-01-2025					
<input type="checkbox"/> Batch do not conform to the requirement and not approved				with date				with date							
Note: Annexure to be provided for additional test report (if required)															