










Quality Inspection Report (Inprocess Inspection)

Report No	QIRI20250852	Select the Appropriate Inspection Stage / Process
Date	18-02-2025.	Inspection of Cut length Wire
Production Order	PR024-2701	NA
Item Code	1.00 CL Wire	
Product Description	Dia 1.00 mm Cut Length Wire 75 mm (Rajaratna)	
DHR No	20250479	
Batch No	WC0018/2025	
Reference Drawing	NA	

Sl. No.	Parameter	Requirement	Instrument Used	Setup Approval (Actual Readings/Observation)										Result (Complied / Not Complied)	Remarks
				1	2	3	4	5	6	7	8	9	10		
1	Diameter	1.00 ± 0.005 mm	DM	1.005	1.005	1.004	1.005	1.004	1.005	1.005	1.005	1.004	1.005	Complied	
2	Length of cut Wires	75 ± 0.50 mm	DVC	74.63	74.69	74.78	74.59	74.62	74.72	74.64	74.73	74.65	74.55	Complied	
3	Straightness	Straight	Rolling Method	OK	OK	OK	OK	OK	OK	OK	OK	OK	OK	Complied	
4	Defects Free	Free from rust, burr, Damage.	Microscope	OK	OK	OK	OK	OK	OK	OK	OK	OK	OK	Complied	
Inspected By (Name)		LAKSHMI						Signature (with date)		 18-02-2025.					

Quality Inspection Report (Inprocess Inspection)

Sl.N o.	Parameter	Date and Time	Instrument Used	Inprocess Samples Inspection (Put "v" - Complied and "X" - Not Complied)										Inspected by (Signature)	Remarks
				1	2	3	4	5	6	7	8	9	10		
1 2 3 4	Diameter Cutting Length Straightness Defects free	18-02-2025	DMM DVC Rolling Method Microscope	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		5
		11:00 AM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		12:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		01:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		02:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		03:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		04:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
Sample Inspected Quantity		200 PCS.	Accepted Batch Quantity		9880 PCS.		Remarks								
Defects Observed (Type)															
Post Quality Inspection Result				Inspected By (Name)		LAKSHMI		Approved By (Name)		MURIGESHAPPA R					
<input checked="" type="checkbox"/> Batch conform to the requirement and approved				Signature				Signature							
<input type="checkbox"/> Batch do not conform to the requirement and not approved				with date		18-02-2025		with date		18-02-2025					
Note: Annexure to be provided for additional test report (if required)															