





Quality Inspection Report (Inprocess Inspection)

Report No	QIRI20254009	Select the Appropriate Inspection Stage / Process
Date	07-08-2025	Inspection of Cut length Wire
Production Order	PR025-1125	
Item Code	Ø 0.90 WIRE	
Product Description	Dia 0.90 mm Cut Length Wire 75 mm	
DHR No	20251870	
Batch No	WC0074/2025	
Reference Drawing	NA	

Sl.N o.	Parameter	Requirement	Instrument Used	Setup Approval (Actual Readings/Observation)										Result (Complied / Not Complied)	Remarks
				1	2	3	4	5	6	7	8	9	10		
1	Diameter	0.90 ± 0.005 mm	DM	0.901	0.900	0.903	0.900	0.903	0.901	0.902	0.903	0.901	0.901	Complied	24
2	Length of cut Wires	75 ± 0.50 mm	DVC	74.54	74.53	74.54	74.53	74.52	74.60	74.53	74.52	74.54	74.63	Complied	
3	Straightness	Straight	Rolling Method	OK	OK	OK	OK	OK	OK	OK	OK	OK	OK	Complied	
4	Defects Free	Free from rust, burr, Damage.	Microscope	OK	OK	OK	OK	OK	OK	OK	OK	OK	OK	Complied	
Inspected By (Name)		LAKSHMI					Signature (with date)					 07-08-2025			

Quality Inspection Report (Inprocess Inspection)

Sl. No.	Parameter	Date and Time	Instrument Used	Inprocess Samples Inspection (Put "v" - Complied and "X" - Not Complied)										Inspected by (Signature)	Remarks
				1	2	3	4	5	6	7	8	9	10		
1	Diameter Cutting Length Straightness Defects free	07-08-2025 03:00 PM	DMM DVC Rolling Method Microscope	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		N/A
		04:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		05:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		06:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		08-08-2025 09:30 AM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		10:30 AM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		11:30 AM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		12:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		01:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		02:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		03:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		04:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		05:50 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		09-08-2025 09:30 AM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		10:30 AM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		11:30 AM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		12:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		01:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		Sample Inspected Quantity		✓	Accepted Batch Quantity		✓		Remarks		✓				
		Defects Observed (Type)		✓											
Post Quality Inspection Result				Inspected By (Name)		LAKSHMI		Approved By (Name)		MURIGESHAPPA R					
<input checked="" type="checkbox"/> Batch conform to the requirement and approved				Signature				Signature							
<input type="checkbox"/> Batch do not conform to the requirement and not approved				with date		09-08-2025		with date							
Note: Annexure to be provided for additional test report (if required)															

Quality Inspection Report (Inprocess Inspection)

Sl. No.	Parameter	Date and Time	Instrument Used	Inprocess Samples Inspection (Put "v" - Complied and "X" - Not Complied)										Inspected by (Signature)	Remarks
				1	2	3	4	5	6	7	8	9	10		
1 2 3 4	Diameter Cutting Length Straightness Defects free	09-08-2025 02:30 PM	DMM DVC Rolling Method Microscope	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	[Signature]	N/A
		03:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		04:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		05:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		06:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		11-08-2025 09:40 AM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		10:40 AM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		11:40 AM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		12:40 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		01:40 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		02:40 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		03:40 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		04:40 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		05:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		06:00 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		Sample Inspected Quantity		Accepted Batch Quantity		Remarks									
Defects Observed (Type)															
Post Quality Inspection Result				Inspected By (Name)		LAKSHMI		Approved By (Name)		MURIGESHAPPA R					
<input checked="" type="checkbox"/> Batch conform to the requirement and approved				Signature		[Signature]		Signature		[Signature]					
<input type="checkbox"/> Batch do not conform to the requirement and not approved				with date		11-08-2025		with date		11-08-2025					
Note: Annexure to be provided for additional test report (if required)															

Quality Inspection Report (Inprocess Inspection)

Sl. No.	Parameter	Date and Time	Instrument Used	Inprocess Samples Inspection (Put "v" - Complied and "X" - Not Complied)										Inspected by (Signature)	Remarks
				1	2	3	4	5	6	7	8	9	10		
1	Diameter	12-08-2025	DMM	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	[Signature]	
		09:30 AM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		10:30 AM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		11:30 AM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		12:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		01:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		02:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		03:30 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
		04:40 PM		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
				✓	✓	✓	✓	✓	✓	✓	✓	✓	✓		
2	Cutting Length	05:30 PM	DVC	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	[Signature]	
3	Straightness		Rolling Method												
4	Defects free		Microscope												
				NOA											
Sample Inspected Quantity		500 PCS.	Accepted Batch Quantity		86700 PCS.					Remarks					
Defects Observed (Type)															
Post Quality Inspection Result				Inspected By (Name)		LAKSHMI		Approved By (Name)		MURIGESHAPPA R					
<input checked="" type="checkbox"/> Batch conform to the requirement and approved				Signature with date		[Signature] 12-08-2025		Signature with date		[Signature] 12-08-2025					
<input type="checkbox"/> Batch do not conform to the requirement and not approved															
Note: Annexure to be provided for additional test report (if required)															