

See discussions, stats, and author profiles for this publication at: <https://www.researchgate.net/publication/221271573>

High-accuracy ultrasound target localization for hand-eye calibration between optical tracking systems and three-dimensional ultrasound

Conference Paper · January 2011

DOI: 10.1007/978-3-642-19335-4_38 · Source: DBLP

CITATIONS

4

READS

189

4 authors, including:



Florian Griesse

University Medical Center Hamburg - Eppendorf

27 PUBLICATIONS 134 CITATIONS

[SEE PROFILE](#)



Floris Ernst

Universität zu Lübeck

161 PUBLICATIONS 1,116 CITATIONS

[SEE PROFILE](#)



Achim Schweikard

Universität zu Lübeck

338 PUBLICATIONS 3,755 CITATIONS

[SEE PROFILE](#)

Some of the authors of this publication are also working on these related projects:



Robotic 4D ultrasound platform for automated imaging and interventional therapy guidance [View project](#)



Nav EVAR [View project](#)

High-accuracy ultrasound target localization for hand-eye calibration between optical tracking systems and three-dimensional ultrasound

Ralf Bruder¹, Florian Griesse², Floris Ernst¹, Achim Schweikard¹

¹Institute for Robotics and Cognitive Systems, University Lübeck

²University Lübeck

bruder@rob.uni-luebeck.de

Abstract. Real-time target localization in ultrasound is useful in many clinical and scientific areas. For example in radiation therapy tumors can be localized in real-time and irradiated with a high accuracy. To measure the position of an ultrasound target in a global coordinate system or to extend the tracking volume by moving the ultrasound transducer an optical marker is attached to it and observed by an optical tracking system. The necessary calibration matrices from marker to ultrasound volume are obtained using hand-eye calibration algorithms which take sets of corresponding observations of the optical marker and an ultrasound target as input. The quality of these calibration matrices is highly dependent on the measured observations. While the accuracy of optical tracking systems is very high, accurate tracking in ultrasound is difficult because of the low resolution of the ultrasound volume, artifacts and noise. Therefore accurate hand-eye calibration is difficult between ultrasound and optical tracking systems. We have tested different phantoms, matching- and sub-pixel strategies to provide highly accurate tracking results in 3D ultrasound volumes as basis for hand-eye calibration. Tests have shown that – using the described methods - calibration results with RMS errors of less than 1mm between observed and calibrated targets can be reached.

1 Introduction

With the recent enhancements in imaging quality and speed three-dimensional ultrasound has become attractive for automatic guidance during robotized interventions and high-accuracy target tracking applications. We use a modified GE Vivid 7 dimension 3D cardiovascular ultrasound station for real-time volume processing and direct target localization in radio-surgery [1]. This station is capable of providing ultrasound volume scans of the target region with more than 20 frames per second. A framework was established to run image-processing algorithms directly on the ultrasound machine in order to track targets inside the ultrasound volume.

To map the obtained target positions to world coordinates the ultrasound transducer is itself tracked using an attached optical marker and an accuTrack


250 (atracsys LLC, Bottens, CH) optical tracking system to locate the marker. The position of the ultrasound target in world coordinates can be calculated using both tracking results and the static translation between ultrasound transducer and optical marker. Among other strategies the best results for this translation are usually obtained [2] using a hand-eye calibration algorithm [3] and the calibration setup shown in figure 1. The equation $T_i \cdot B \cdot U_i = C$ is solved for the unknown calibration matrix B by capturing multiple sets of corresponding optical and ultrasound positions T_i and U_i for different transducer positions.

One problem of hand-eye calibration is that the quality of the measured position sets highly influences the calibration result. While the measurements of optical tracking systems are highly accurate, tracking in ultrasound is difficult because of the low resolution of the volume data, artifacts and noise. To obtain an accurate calibration this tracking has to be optimized.

2 Materials and Methods

A calibration setup is used (Fig. 1). An application has been developed to capture both ultrasound and optical tracking results. To obtain corresponding datasets from both systems the capture time of both measurements is calculated using high-accuracy timestamps and an estimate for the system latencies. Then the optical tracking result is chosen to meet the capture time of the corresponding ultrasound volume. For an intuitive handling during hand-held sample acquisition an automatic capture trigger algorithm has been implemented besides the manually triggered capture process. This algorithm triggers in phases with minimal movements in both tracking results. In this way motion artifacts in the measured tracking results are further minimized. After completion of a predefined sample count the result of the hand-eye calibration is calculated and the distance error between the calibration and the measured points is computed. Depending on these values an iterative post-processing algorithm identifies up to

Fig. 1. Setup for hand-eye calibration



ten percent of the samples with high distances, eliminates these samples and recalculates the tracking result. In this way errors such as wrong target detections in ultrasound are eliminated.

Two types of phantoms are used for tracking in ultrasound: A single target (Figure 1a) is used to maximize tracking accuracy while later on a complex phantom (Fig. 1b) with multiple features is used to increase targeting accuracy and reduce the amount of necessary count of measured transducer positions.

2.1 Tracking a single target

A lead bulb on a nylon wire in a liquid tank is used as target. The lead bulb reflects a high amount of ultrasound and can be easily detected in the ultrasound volume using a maximum intensity search. As our ultrasound volume has a worst-case resolution of more than 1.5mm per pixel this method is not sufficient for high-accuracy tracking. To overcome this limitation we use cubic splines to interpolate the target region around a detected position with maximum intensity. The extreme value of the interpolated volume is used as sub-pixel approximation of the lead bulb position. Another tracking possibility is template matching. Using a predefined pattern of the lead bulb, the sum of square distances (SSD) is calculated to find an optimal match between the pattern and the lead bulb. With sub-pixel enhancement through iterative interpolation of the target region a high accuracy tracking result can be obtained.

Figure 3b shows the different tracking positions for a two-dimensional slice of the target. To test the sub-pixel quality of each strategy a lead bulb is moved through a liquid tank on a linear trajectory while being continuously tracked (Figure 3c). While the tracking accuracy is increased with both strategies, template matching shows the best linearity and highest accuracy.

2.2 Tracking multiple targets

Three-dimensional ultrasound offers the possibility to track multiple targets inside one volume simultaneously. We use a complex wire phantom (Dansk Fantom




Fig. 2. Ultrasound transducer with attached optical marker on phantoms. Left: lead bulb phantom; right: complex phantom.

Service, Fig. 4a) to find a defined geometry of twisted nylon wires inside the volume. The wires are clearly visible in the ultrasound volume and can be easily tracked using a repeated maximum intensity search with an additional distance criterium, so that features in close proximity to already located points are not detected. Alternatively template matching can be used with multiple, predefined patterns for wires in different angles to the beam.

The obtained feature positions are matched against a predefined virtual phantom or the first captured phantom using the ICP algorithm [4] which may be exchanged replaced by RANSAC [5] to eliminate erroneous feature detections. The resulting transformation matrix between virtual and detected feature cloud is used as tracking result U_i . Figure 4 shows the alignment results.




Fig. 3. A captured lead bulb with artifacts is shown in (a). The crosses in (b) show the maximum intensity (red), the spline-interpolated maximum intensity (black) and the center position of a found template using sub-pixel matching (blue). (c) shows the observed position changes for a linear lead bulb movement using the same color scheme.




Fig. 4. The complex wire phantom in ultrasound is shown in (a). In (b) the extracted feature cloud (green) is registered to the predefined phantom (blue). While RANSAC (black) works as expected the ICP result (red) is rotated due to false feature detections.

Table 1. Calibration results.

| Calibration method | Positions | RMS error |
|--------------------------------------|------------------------------|-----------|
| Lead bulb, standard resolution | 100 | 1.811mm |
| Lead bulb, sub-pixel enhanced | 50 | 0.773mm |
| Complex phantom, standard resolution | 3 positions, 50 targets each | 2.330mm |
| Complex phantom, sub-pixel enhanced | 3 positions, 50 targets each | 0.505mm |

3 Results

To measure the calibration quality we compute the RMS distance error between all calibrated and measured target points. Table 1 shows this quality value for the different calibration scenarios. For the single target an automatic calibration with 100 captured points is performed without manual interaction. Measurements at three different positions with 50 located targets in each volume are performed for the complex phantom. Both calibration methods show acceptable results with RMS errors of less than 2mm with standard resolution. With high-accuracy ultrasound target localization both methods can be significantly enhanced. For single targets the RMS errors can be reduced to less than 1mm using only half of captured points. As ICP is very sensitive to position errors, the accuracy gain for complex phantom calibration due to better target localization is even higher.

4 Conclusion

Two calibration methods are implemented and enhanced with sub-pixel target localization. Especially with complex phantoms a stable calibration can be performed in only three steps which makes the method attractive for common use.

References

1. Bruder R, Ernst F, Schlaefer A, et al. Real-time tracking of the pulmonary veins in 3D ultrasound of the beating heart. In: 51st Annual Meeting of the AAPM. vol. 36 of Med Phys; 2009. p. 2804.
2. Poon TC, Rohling RN. Comparison of calibration methods for spatial tracking of a 3-D ultrasound probe. Eur J Ultrasound. 2005;31(8):1095–108.
3. Tsai RY, Lenz RK. A new technique for fully autonomous and efficient 3D robotics hand/eye calibration. IEEE Trans Rob Autom. 1989;5(3):345–58.
4. Rusinkiewicz S, Levoy M. Efficient variants of the ICP algorithm. In: Proc Int Conf 3D Digit Imaging Model; 2001. p. 145–52.
5. Chen C, Hung Y, Cheng J. RANSAC-based DARCES: a new approach to fast automatic registration of partially overlapping range images. IEEE Trans Pattern Anal Mach Intell. 1999;21:1229–34.