

**Standard Name** 8 of 11

9B 0860 CRD

Visual Hue Request:

New Std	Retrieve Std	Store Std	List Std	Update Std
---------	--------------	-----------	----------	------------

Instrument Setup: %R LAV SCI UV Cal

Date: 07-Oct-19      Time: 08:58:11

**Batch Name** 3 of 3

↑
SJS7091 PR2122/68716 1-1 3052636/VII/21
↓

New Bat	Retrieve Bat	Store Bat	List Bat	Update Bat
---------	--------------	-----------	----------	------------

Instrument Setup: %R LAV SCI UV Cal

Date: 02-Oct-21      Time: 09:11:50

**CMC DE Pass/Fail Tolerance: 1.00**

I:c Ratio: ( 2.00 1.00)

Measured by: ▼

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/F	Color	Comments	
SJS7091 PR2122/68716 1-1 3052636/VII/21	D65 10 Deg	0.38	0.29	0.69	0.18	0.34	-0.37	<b>0.54</b>		Pass	too light	too bright	too yellow
SJS7091 PR2122/68716 1-1 3052636/VII/21	F11 10 Deg	0.82	-0.43	1.11	0.39	0.63	0.30	<b>0.80</b>	0.96	Pass	too light	too bright	too green
SJS7091 PR2122/68716 1-1 3052636/VII/21	A 10 Deg	0.42	-1.47	1.35	0.20	0.69	1.79	<b>1.93</b>	1.92	Warn	too light	too bright	too green

DataColor TOOLS 2.8.1

**S/N NO. 8823242**

