

Standard Name 10 of 62

↑

21B 0347-2 ARD

↓

Visual Hue Request:

New Std

Retrieve Std

Store Std

List Std

Update Std

Instrument Setup:%R LAV SCI UV Cal

Date: 07-May-21 Time: 09:10:33

Batch Name 1 of 2

↑

MTR17080 B22013010AA-003 4-1 3053176/VII/21

↓

New Bat

Retrieve Bat

Store Bat

List Bat

Update Bat

Instrument Setup: %R LAV SCI UV Cal

Date: 09-Sep-21 Time: 11:16:26

CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: ( 2.00 1.00)


Measured by:

Delta 

D65 / 10

F11 / 10

A / 10



- Single/All
- CMC Plot
- Zoom Out
- Zoom In
- Print Form

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/E	Color	Comments	
MTR17080 B22013010AA-003 4-1 3053176/VII/21.	D65 10 Deg	0.27	0.23	-0.79	0.14	0.41	0.43	0.61		Pass	too light	too bright	too red
MTR17080 B22013010AA-003 4-1 3053176/VII/21.	F11 10 Deg	0.23	0.09	-0.87	0.12	0.46	0.26	0.54	0.15	Pass	too light	too bright	too red
MTR17080 B22013010AA-003 4-1 3053176/VII/21.	A 10 Deg	0.22	-0.03	-0.85	0.12	0.43	0.25	0.51	0.23	Pass	too light	too bright	too red