

Standard Name 7 of 7

↑

20B 0484-2 BRD

↓

Visual Hue Request:

New Std	Retrieve Std	Store Std	List Std	Update Std
---------	--------------	-----------	----------	------------

Instrument Setup:%R LAV SCI UV Cal

Date: 29-Jul-21 Time: 14:33:15

Batch Name 1 of 2

↑

MTR17080 BOOKING FA22 13-1 3088202/XII/21

↓

New Bat	Retrieve Bat	Store Bat	List Bat	Update Bat
---------	--------------	-----------	----------	------------

Instrument Setup: %R LAV SCI UV Cal

Date: 07-Jan-22 Time: 14:52:10

CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: (2.00 1.00)


Measured by:

Delta

D65 / 10

F11 / 10

A / 10



Single/All

CMC Plot

Zoom Out

Zoom In

Print Form

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/F	Color	Comments	
MTR17080 BOOKING FA22 13-1 3088202/XII/21	D65 10 Deg	-1.03	-0.13	0.52	-0.49	0.10	-0.58	0.76		Warn	too dark	too bright	too yellow
MTR17080 BOOKING FA22 13-1 3088202/XII/21	F11 10 Deg	-0.99	-0.13	0.59	-0.47	0.05	-0.68	0.83	0.08	Pass	too dark	too bright	too yellow
MTR17080 BOOKING FA22 13-1 3088202/XII/21	A 10 Deg	-1.00	-0.01	0.53	-0.48	-0.19	-0.57	0.77	0.16	Pass	too dark	too dull	too yellow