

Standard Name 67 of 76

↑

20B 0360-1 BRM

↓

Visual Hue Request:

New Std

Retrieve Std

Store Std

List Std

Update Std

Instrument Setup:%R LAV SCI UV Cal

Date: 18-Jun-20 Time: 10:13:48

Batch Name 1 of 2

↑

R17080 B22173810AA-004 1-1 3046878/VIII/22

↓

New Bat

Retrieve Bat

Store Bat

List Bat

Update Bat

Instrument Setup: %R LAV SCI UV Cal


Date: 24-Aug-22 Time: 08:12:39

CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: (2.00 1.00)

Measured by:

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/E	Color Comments		
MTR17080 B22173810AA-004 1-1 3046878/VIII/22	D65 10 Deg	-1.68	0.43	0.22	-0.67	0.16	-0.59	0.91		Warn	too dark	too bright	too yellow
MTR17080 B22173810AA-004 1-1 3046878/VIII/22	F11 10 Deg	-1.57	-0.14	0.01	-0.63	0.00	0.19	0.66	0.59	Pass	too dark	too bright	too green
MTR17080 B22173810AA-004 1-1 3046878/VIII/22	A 10 Deg	-1.62	-0.18	0.67	-0.64	0.52	0.65	1.06	0.71	Pass	too dark	too bright	too yellow



Single/All

CMC Plot

Zoom Out

Zoom In

Print Form