

Standard Name 225 of 229

↑

21B 0033-1 BRM

↓

Visual Hue Request:

New Std	Retrieve Std	Store Std	List Std	Update Std
---------	--------------	-----------	----------	------------

Instrument Setup:%R LAV SCI UV Cal

Date: 01-Mar-21 Time: 16:05:29

Batch Name 1 of 3

↑

F3C17109 PR2122/68792 22-7 3052704/VII/21

↓

New Bat	Retrieve Bat	Store Bat	List Bat	Update Bat
---------	--------------	-----------	----------	------------

Instrument Setup: %R LAV SCI UV Cal

Date: 20-Sep-21 Time: 08:46:57

CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: (2.00 1.00)


Measured by:

Delta

D65 / 10

F11 / 10

A / 10



- Single/All
- CMC Plot
- Zoom Out
- Zoom In
- Print Form

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/F	Color	Comments	
F3C17109 PR2122/68792 22-7 3052704/VII/21	D65 10 Deg	0.03	0.10	0.77	0.01	-0.11	-0.90	0.91		Warn	too light	too dull	too yellow
F3C17109 PR2122/68792 22-7 3052704/VII/21	F11 10 Deg	0.06	0.18	0.85	0.02	-0.24	-0.97	1.00	0.12	Pass	too light	too dull	too yellow
F3C17109 PR2122/68792 22-7 3052704/VII/21	A 10 Deg	0.10	0.32	0.80	0.04	-0.53	-0.80	0.96	0.23	Pass	too light	too dull	too yellow