

Standard Name 145 of 192

↑

21C 0886 BRL

↓

Visual Hue Request:

New Std	Retrieve Std	Store Std	List Std	Update Std
---------	--------------	-----------	----------	------------

Instrument Setup:%R LAV SCI UV Cal

Date: 14-Oct-21 Time: 13:58:48

Batch Name 3 of 3

↑

R1117235 B22147210AA-001 1-1 3042057/VII/22

↓

New Bat	Retrieve Bat	Store Bat	List Bat	Update Bat
---------	--------------	-----------	----------	------------

Instrument Setup: %R LAV SCI UV Cal

Date: 01-Sep-22 Time: 11:09:28


CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: (2.00 1.00)

Measured by:

▼

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/E	Color Comments		
R1117235 B22147210AA-001 1-1 3042057/VII/22	D65 10 Deg	-1.12	0.14	0.52	-0.40	0.52	-0.28	0.72		Pass	too dark	too bright	too yellow
R1117235 B22147210AA-001 1-1 3042057/VII/22	F11 10 Deg	-1.09	0.20	0.54	-0.39	0.56	-0.24	0.73	0.08	Pass	too dark	too bright	too red
R1117235 B22147210AA-001 1-1 3042057/VII/22	A 10 Deg	-1.06	0.28	0.58	-0.38	0.69	-0.11	0.80	0.18	Warn	too dark	too bright	too red



Single/All

CMC Plot

Zoom Out

Zoom In

Print Form