

Standard Name 77 of 82

↑

21B 0030-2 DRD

↓

Visual Hue Request:

New Std

Retrieve Std

Store Std

List Std

Update Std

Instrument Setup:%R LAV SCI UV Cal

Date: 27-Apr-21 Time: 13:59:57

Batch Name 1 of 2

↑

MTR17080 B22013010AA-003 8-1 3053176/VII/21

↓

New Bat

Retrieve Bat

Store Bat

List Bat

Update Bat

Instrument Setup: %R LAV SCI UV Cal


Date: 10-Sep-21 Time: 11:36:46

CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: (2.00 1.00)

Measured by:

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/E	Color	Comments	
MTR17080 B22013010AA-003 8-1 3053176/VII/21	D65 10 Deg	-0.15	0.12	0.52	-0.08	0.26	-0.12	0.30		Pass	too dark	too bright	too red
MTR17080 B22013010AA-003 8-1 3053176/VII/21	F11 10 Deg	-0.22	0.21	0.24	-0.11	0.09	-0.18	0.23	0.28	Pass	too dark	too bright	too yellow
MTR17080 B22013010AA-003 8-1 3053176/VII/21	A 10 Deg	-0.16	-1.21	0.68	-0.08	0.35	1.01	1.08	1.34	Pass	too dark	too bright	too green



- Single/All
- CMC Plot
- Zoom Out
- Zoom In
- Print Form