

Standard Name 171 of 171

9C 1166 BRD

Visual Hue Request:

New Std	Retrieve Std	Store Std	List Std	Update Std
---------	--------------	-----------	----------	------------

Instrument Setup: %R LAV SCI UV Cal

Date: 24-Dec-19 Time: 09:40:13

Batch Name 3 of 3

↑ SJS7091 572584 1-1 3052261/VII/21 ↓

New Bat	Retrieve Bat	Store Bat	List Bat	Update Bat
---------	--------------	-----------	----------	------------

Instrument Setup: %R LAV SCI UV Cal

Date: 15-Sep-21 Time: 13:44:16

CMC DE Pass/Fail Tolerance: 1.00


I:c Ratio: (2.00 1.00)

Measured by: ▼

<u>Batch Name</u>	<u>III/Obs</u>	<u>DL*</u>	<u>Da*</u>	<u>Db*</u>	<u>DL*/ISL</u>	<u>DC*/cSC</u>	<u>DH*/SH</u>	<u>CMC DE</u>	<u>M.I.</u>	<u>P/F</u>	<u>Color</u>	<u>Comments</u>	
SJS7091 572584 1-1 3052261/VII/21	D65 10 Deg	1.71	-0.35	-0.93	0.81	-0.25	-0.42	0.95		Warn	too light	too dull	too blue
SJS7091 572584 1-1 3052261/VII/21	F11 10 Deg	1.60	0.02	-1.37	0.72	-0.25	-0.77	1.08	0.84	Pass	too light	too dull	too blue
SJS7091 572584 1-1 3052261/VII/21	A 10 Deg	1.61	-0.81	-0.83	0.70	-0.37	-0.09	0.80	0.55	Pass	too light	too dull	too blue

DataColor TOOLS 2.8.1

S/N NO. 8823242



Single/All

CMC Plot

Zoom Out

Zoom In

Print Form