

Standard Name 19 of 19

↑

21C 0160-1 ARD

↓

Visual Hue Request:

New Std	Retrieve Std	Store Std	List Std	Update Std
---------	--------------	-----------	----------	------------

Instrument Setup:%R LAV SCI UV Cal  
Date: 01-Jul-21 Time: 15:55:18

Batch Name 1 of 2

↑

F3T18157 B21228101AA-002 6-1 3075826/X/21

↓

New Bat	Retrieve Bat	Store Bat	List Bat	Update Bat
---------	--------------	-----------	----------	------------


Instrument Setup: %R LAV SCI UV Cal  
Date: 23-Nov-21 Time: 09:04:14

CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: ( 2.00 1.00)

Measured by:

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/E	Color	Comments	
F3T18157 B21228101AA-002 6-1 3075826/X/21	D65 10 Deg	1.16	-0.66	-0.03	0.62	-0.39	0.49	0.88		Warn	too light	too dull	too yellow
F3T18157 B21228101AA-002 6-1 3075826/X/21	F11 10 Deg	1.10	-0.57	-0.16	0.58	-0.36	0.37	0.78	0.21	Pass	too light	too dull	too yellow
F3T18157 B21228101AA-002 6-1 3075826/X/21	A 10 Deg	1.07	-0.79	-0.19	0.55	-0.46	0.47	0.86	0.29	Pass	too light	too dull	too yellow



- Single/All
- CMC Plot
- Zoom Out
- Zoom In
- Print Form