

Standard Name    171   of   171

↑

9C 1166 BRD

↓

Visual Hue Request:

New Std	Retrieve Std	Store Std	List Std	Update Std
---------	--------------	-----------	----------	------------

Instrument Setup:%R LAV  SCI UV Cal

Date:  24-Dec-19                      Time:  09:40:13

Batch Name                      2   of   3

↑

R1117235 572584 1-1 3052261/VII/21

↓

New Bat	Retrieve Bat	Store Bat	List Bat	Update Bat
---------	--------------	-----------	----------	------------

Instrument Setup: %R LAV  SCI UV Cal

Date:  15-Sep-21                      Time:  13:43:37

CMC DE Pass/Fail Tolerance:    1.00

I:c Ratio: (   2.00   1.00 )

Measured by: 

▼

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/F	Color Comments		
R1117235 572584 1-1 3052261/VII/21	D65 10 Deg	0.47	-0.85	-0.57	0.23	-0.34	-0.07	<b>0.42</b>		Pass	too light	too dull	too blue
R1117235 572584 1-1 3052261/VII/21	F11 10 Deg	0.20	-0.86	-1.29	0.09	-0.48	-0.40	<b>0.63</b>	0.82	Pass	too light	too dull	too blue
R1117235 572584 1-1 3052261/VII/21	A 10 Deg	0.35	-1.03	-0.61	0.15	-0.38	0.13	<b>0.43</b>	0.27	Pass	too light	too dull	too yellow

