

Standard Name 225 of 229

↑

21B 0033-1 BRM

↓

Visual Hue Request:

New Std	Retrieve Std	Store Std	List Std	Update Std
---------	--------------	-----------	----------	------------

Instrument Setup:%R LAV SCI UV Cal

Date: 01-Mar-21 Time: 16:05:29

Batch Name 2 of 3

↑

SJS7091 PR2122/68792 1-1 3052704/VII/21 CC

↓

New Bat	Retrieve Bat	Store Bat	List Bat	Update Bat
---------	--------------	-----------	----------	------------

Instrument Setup: %R LAV SCI UV Cal

Date: 20-Sep-21 Time: 08:47:14

CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: (2.00 1.00)


Measured by:

Delta

D65 / 10

F11 / 10

A / 10



- Single/All
- CMC Plot
- Zoom Out
- Zoom In
- Print Form

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/F	Color	Comments	
SJS7091 PR2122/68792 1-1 3052704/VII/21 CCK F3C171	D65 10 Deg	-1.29	-0.24	0.65	-0.52	0.22	-0.77	0.96		Warn	too dark	too bright	too yellow
SJS7091 PR2122/68792 1-1 3052704/VII/21 CCK F3C171	F11 10 Deg	-1.24	-0.43	0.72	-0.50	0.34	-0.88	1.07	0.21	Pass	too dark	too bright	too yellow
SJS7091 PR2122/68792 1-1 3052704/VII/21 CCK F3C171	A 10 Deg	-1.26	-0.11	0.66	-0.52	-0.09	-0.79	0.95	0.15	Pass	too dark	too dull	too yellow