

Standard Name 14 of 14

↑

21B 0452 CRD

↓

Visual Hue Request:

New Std

Retrieve Std

Store Std

List Std

Update Std

Instrument Setup:%R LAV SCI UV Cal

Date: 10-May-21 Time: 14:33:50

Batch Name 1 of 2

↑

MTR17080 B21243501AA-003 1-1 3075994/X/21

↓

New Bat

Retrieve Bat

Store Bat

List Bat

Update Bat

Instrument Setup: %R LAV SCI UV Cal

Date: 22-Nov-21 Time: 09:04:53

CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: ( 2.00 1.00)


Measured by:

Delta 

D65 / 10

F11 / 10

A / 10



Single/All

CMC Plot

Zoom Out

Zoom In

Print Form

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/F	Color	Comments	
MTR17080 B21243501AA-003 1-1 3075994/X/21	D65 10 Deg	-0.05	0.01	0.34	-0.02	-0.13	-0.22	0.26		Pass	too dark	too dull	too yellow
MTR17080 B21243501AA-003 1-1 3075994/X/21	F11 10 Deg	-0.02	0.11	0.40	-0.01	-0.22	-0.17	0.28	0.12	Pass	too dark	too dull	too yellow
MTR17080 B21243501AA-003 1-1 3075994/X/21	A 10 Deg	-0.02	-0.00	0.36	-0.01	-0.14	-0.20	0.24	0.04	Pass	too dark	too dull	too yellow