

Standard Name 30 of 33

↑

21B 0335 ARL

↓

Visual Hue Request:

New Std

Retrieve Std

Store Std

List Std

Update Std

Instrument Setup:%R LAV SCI UV Cal

Date: 15-Apr-21 Time: 11:27:34

Batch Name 2 of 2

↑

F3C17109 PR2122/68792 5-1 3052704/VII/21

↓

New Bat

Retrieve Bat

Store Bat

List Bat

Update Bat

Instrument Setup: %R LAV SCI UV Cal

Date: 03-Oct-21 Time: 09:30:40

CMC DE Pass/Fail Tolerance: 1.00

I:c Ratio: (2.00 1.00)


Measured by:

Delta

D65 / 10

F11 / 10

A / 10



- Single/All
- CMC Plot
- Zoom Out
- Zoom In
- Print Form

Batch Name	III/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/F	Color	Comments	
F3C17109 PR2122/68792 5-1 3052704/VII/21	D65 10 Deg	-0.67	0.50	1.16	-0.25	0.35	-0.90	1.00		Warn	too dark	too bright	too yellow
F3C17109 PR2122/68792 5-1 3052704/VII/21	F11 10 Deg	-0.60	0.49	1.45	-0.22	0.51	-0.99	1.14	0.32	Warn	too dark	too bright	too yellow
F3C17109 PR2122/68792 5-1 3052704/VII/21	A 10 Deg	-0.55	0.75	1.20	-0.21	0.35	-1.17	1.23	0.29	Pass	too dark	too bright	too yellow