

Standard Name25 of 55

AE6S Wonder WhiteF21

Visual Hue Request:

New StdRetrieve StdStore StdList Std

Instrument Setup:%R LAV SCI UV Cal

Date: 14-Dec-19Time: 07:56:49

Batch Name1 of 1

733214 POAO2104A/E0644 1-1 3027336/I

New BatRetrieve BatStore BatList Bat


Instrument Setup: %R LAV SCI UV Cal

Date: 10-May-21Time: 19:04:43

Batch Name

62733214 POAO2104A/E0644 1-1 3027336/IV/21

62733214 POAO2104A/E0644 1-1 3027336/IV/21



Single/AllZoom InZoom OutCMC PlotPrint Form

DataColor TOOLS 2.4.4S/N NO. 8823242

Standard %R

400-450nm64.6765.5262.7959.0558.9555.65

460-510nm58.6467.3071.7272.2271.9770.54

520-570nm68.7769.2669.9869.7669.5171.94

580-630nm75.5677.8479.1579.8079.5178.71

640-690nm78.7379.6983.0987.1790.6591.89

700nm92.96

CMC DE Pass/Fail Tolerance: 0.80

I:c Ratio: (2.00 1.00)

Ill/Obs	DL*	Da*	Db*	DL*/ISL	DC*/cSC	DH*/SH	CMC DE	M.I.	P/F	Color	Comments
D65 10 Deg	-0.78	0.38	0.36	-0.28	0.36	-0.45	0.64		Warn	darker	more saturated more red
F11 10 Deg	-0.67	-0.62	0.27	-0.24	0.08	1.03	1.06	1.01	Fail	darker	more saturated more yellow